

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE101918\
 Data File : BE097991.D
 Acq On : 19 Oct 2018 14:48
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 19 23:49:11 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE101618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 16 15:36:44 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	135	0.00
2	1,4-Dioxane	0.400	0.358	10.5	131	0.00
3	n-Nitrosodimethylamine	0.400	0.327	18.3	125	0.00
4 S	2-Fluorophenol	0.400	0.389	2.8	139	0.00
5 S	Phenol-d6	0.400	0.411	-2.7	142	0.00
6	bis(2-Chloroethyl)ether	0.400	0.362	9.5	130	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	155	0.00
8 S	Nitrobenzene-d5	0.400	0.344	14.0	136	0.00
9	Naphthalene	0.400	0.379	5.3	151	0.00
10	Hexachlorobutadiene	0.400	0.370	7.5	148	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.391	2.3	155	0.00
12	2-Methylnaphthalene	0.400	0.398	0.5	159	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	169	0.00
14 S	2,4,6-Tribromophenol	0.400	0.323	19.3	154	0.00
15 S	2-Fluorobiphenyl	0.400	0.341	14.8	152	0.00
16	Acenaphthylene	0.400	0.362	9.5	161	0.00
17	Acenaphthene	0.400	0.368	8.0	163	0.00
18	Fluorene	0.400	0.350	12.5	155	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	159	0.00
20	4-Bromophenyl-phenylether	0.400	0.378	5.5	154	0.00
21	Hexachlorobenzene	0.400	0.391	2.3	159	0.00
22	Pentachlorophenol	0.400	0.472	-18.0	214	-0.01
23	Phenanthrene	0.400	0.358	10.5	147	0.00
24	Anthracene	0.400	0.357	10.8	149	0.00
25 SURR	Fluoranthene-d10	0.400	0.325	18.8	139	0.00
26	Fluoranthene	0.400	0.327	18.3	139	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	145	-0.01
28	Pvrene	0.400	0.358	10.5	138	0.00
29 S	Terphenyl-d14	0.400	0.341	14.8	137	0.00
30	Benzo(a)anthracene	0.400	0.353	11.8	137	0.00
31	Chrysene	0.400	0.366	8.5	146	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.339	15.3	132	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.327	18.3	131	-0.01
34 I	Perylene-d12	0.400	0.400	0.0	143	-0.01
35	Benzo(b)fluoranthene	0.400	0.356	11.0	137	0.00
36	Benzo(k)fluoranthene	0.400	0.353	11.8	139	0.00
37 C	Benzo(a)pyrene	0.400	0.350	12.5	133	-0.01
38	Dibenzo(a,h)anthracene	0.400	0.349	12.8	133	0.00
39	Benzo(a,h,i)perylene	0.400	0.342	14.5	134	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE101918\
Data File : BE097991.D
Acq On : 19 Oct 2018 14:48
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 19 23:49:11 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE101618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 16 15:36:44 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				