

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE101918\
 Data File : BE098001.D
 Acq On : 19 Oct 2018 20:56
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 19 23:21:07 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE101618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 16 15:36:44 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	-0.01
2	1,4-Dioxane	0.697	0.761	-9.2	120	0.00
3	n-Nitrosodimethylamine	0.779	0.743	4.6	109	0.00
4 S	2-Fluorophenol	1.261	1.355	-7.5	115	-0.01
5 S	Phenol-d6	1.864	2.222	-19.2	124	-0.01
6	bis(2-Chloroethyl)ether	1.583	1.604	-1.3	110	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	118	-0.01
8 S	Nitrobenzene-d5	0.379	0.398	-5.0	127	-0.01
9	Naphthalene	3.983	4.323	-8.5	132	-0.01
10	Hexachlorobutadiene	0.215	0.241	-12.1	137	-0.01
11 SURR	2-Methylnaphthalene-d10	0.665	0.768	-15.5	140	-0.02
12	2-Methylnaphthalene	0.682	0.773	-13.3	139	-0.02
13 I	Acenaphthene-d10	1.000	1.000	0.0	129	-0.02
14 S	2,4,6-Tribromophenol	0.215	0.209	2.8	142	-0.01
15 S	2-Fluorobiphenyl	1.635	1.639	-0.2	137	-0.02
16	Acenaphthylene	1.848	1.942	-5.1	143	0.00
17	Acenaphthene	1.129	1.187	-5.1	142	0.00
18	Fluorene	1.554	1.586	-2.1	138	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	122	-0.01
20	4-Bromophenyl-phenylether	0.219	0.242	-10.5	139	-0.01
21	Hexachlorobenzene	0.215	0.249	-15.8	145	0.00
22	Pentachlorophenol	0.062	0.082	-32.3#	214#	-0.01
23	Phenanthrene	1.016	1.045	-2.9	130	-0.01
24	Anthracene	0.960	0.990	-3.1	132	-0.01
25 SURR	Fluoranthene-d10	5.421	5.205	4.0	126	-0.01
26	Fluoranthene	1.334	1.279	4.1	126	-0.01
27 I	Chrysene-d12	1.000	1.000	0.0	111	-0.01
28	Pvrene	1.154	1.211	-4.9	124	-0.01
29 S	Terphenyl-d14	0.917	0.951	-3.7	127	-0.01
30	Benzo(a)anthracene	1.227	1.282	-4.5	123	0.00
31	Chrysene	1.162	1.260	-8.4	131	-0.01
32	Bis(2-ethylhexyl)phthalate	2.865	2.838	0.9	113	-0.01
33	Indeno(1,2,3-cd)pyrene	1.249	1.342	-7.4	120	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	109	-0.02
35	Benzo(b)fluoranthene	1.151	1.202	-4.4	122	-0.01
36	Benzo(k)fluoranthene	1.199	1.282	-6.9	128	-0.01
37 C	Benzo(a)pyrene	1.122	1.193	-6.3	123	-0.01
38	Dibenzo(a,h)anthracene	1.099	1.138	-3.5	120	-0.02
39	Benzo(a,h,i)perylene	1.126	1.142	-1.4	121	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE101918\
Data File : BE098001.D
Acq On : 19 Oct 2018 20:56
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 19 23:21:07 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE101618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 16 15:36:44 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			