


Data Path : Z:\HPCHEM1\BNA E\Data\BE102014\
 Data File : BE088124.D
 Acq On : 21 Oct 2014 3:27
 Operator : TP/JJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-PAH-BE102014.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Oct 21 02:11:14 2014



AutoFind: Scans 2646, 2647, 2648; Background Corrected with Scan 2639

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	53.2	15443	PASS
68	69	0.00	2	1.2	162	PASS
69	198	0.00	100	48.4	14056	PASS
70	69	0.00	2	0.6	81	PASS
127	198	10	80	54.1	15717	PASS
197	198	0.00	2	0.2	47	PASS
198	198	100	100	100.0	29030	PASS
199	198	5	9	6.4	1853	PASS
275	198	10	60	22.0	6398	PASS
365	198	1	100	2.6	762	PASS
441	442	0.01	24	7.1	1972	PASS
442	198	50	100	95.9	27851	PASS
443	442	15	24	19.7	5500	PASS