

Data Path : Z:\HPCHEM1\BNA E\DATA\BE102017\
 Data File : BE094480.D
 Acq On : 20 Oct 2017 18:27
 Operator : SJ/JU
 Sample : I5840-09
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :

Quant Time: Oct 22 02:47:52 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE092017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 18 17:51:02 2017
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.90	152	1436	0.40	ng/ul	0.00
2) Naphthalene-d8	10.69	136	7934	0.40	ng/ul	-0.01
6) Acenaphthene-d10	14.51	164	4994	0.40	ng/ul	0.00
10) Phenanthrene-d10	0.00	188	0	0.00	ng/ul	-17.26
16) Chrysene-d12	21.42	240	10171	0.40	ng/ul	0.00
20) Perylene-d12	23.78	264	874178	0.40	ng/ul	-0.17
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.28	152	4693	0.38	ng/ul	0.00
14) Fluoranthene-d10	19.28	212	12667	0.00	ng/ul	0.00
Target Compounds						
17) Pyrene	19.64	202	1815	0.064	ng/ul#	Ovalue 1


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE102017\
Data File : BE094480.D
Acq On : 20 Oct 2017 18:27
Operator : SJ/JU
Sample : I5840-09
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :

Quant Time: Oct 22 02:47:52 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE092017.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 18 17:51:02 2017
Response via : Initial Calibration





#17

Pvrene

Concen: 0.064 ng/ul

RT: 19.64 min Scan# 2825

Delta R.T. -0.04 min


Lab File: BE094480.D

Acq: 20 Oct 2017 18:27

Instrument :

BNA_E

ClientSampleId :



Tot Ion:	202	Resp:	1815
Ion Ratio	Lower	Upper	
202	100		
101	157.6	18.2	27.4#
100	813.9	15.6	23.4#

