

Data Path : Z:\HPCHEM1\BNA_E\Data\BE102114\
 Data File : Be088141.d
 Acq On : 21 Oct 2014 14:10
 Operator : TP/JJ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Oct 22 04:36:16 2014
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-PAH-BE102014.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 21 02:11:14 2014
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	0.00
2	1,4-Dioxane	0.719	0.748	-4.0	97	-0.01
3 I	Naphthalene-d8	1.000	1.000	0.0	98	0.00
4 S	Nitrobenzene-d5	0.360	0.361	-0.3	94	0.00
5	Naphthalene	1.031	1.071	-3.9	98	0.00
6	2-Methylnaphthalene	0.617	0.642	-4.1	97	0.00
7 I	Acenaphthene-d10	1.000	1.000	0.0	97	0.00
8 S	2-Fluorobiphenyl	1.381	1.449	-4.9	98	0.00
9	Acenaphthylene	7.619	7.981	-4.8	97	0.00
10	Acenaphthene	1.176	1.219	-3.7	100	0.00
11	Fluorene	1.246	1.293	-3.8	97	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	98	0.00
13	Phenanthrene	4.779	4.099	14.2	79	0.00
14	Anthracene	3.902	3.898	0.1	95	0.00
15	Fluoranthene	0.951	0.925	2.7	95	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	96	0.00
17	Pyrene	1.359	1.354	0.4	96	0.00
18 S	Terphenyl-d14	0.805	0.804	0.1	94	0.00
19	Benzo(a)anthracene	4.447	4.177	6.1	92	0.00
20	Chrysene	4.363	4.477	-2.6	98	0.00
21	Indeno(1,2,3-cd)pyrene	4.846	4.248	12.3	84	0.00
22 I	Perylene-d12	1.000	1.000	0.0	93	0.00
23	Benzo(b)fluoranthene	1.190	1.094	8.1	85	0.00
24	Benzo(k)fluoranthene	1.274	1.287	-1.0	93	0.00
25 C	Benzo(a)pyrene	1.096	1.009	7.9	83	0.00
26	Dibenzo(a,h)anthracene	1.075	0.961	10.6	83	0.00
27	Benzo(g,h,i)perylene	4.752	4.458	6.2	86	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0