

Data Path : Z:\HPCHEM1\BNA_E\Data\Be102116\
 Data File : BE092464.D
 Acq On : 21 Oct 2016 11:51
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 21 17:26:19 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE092216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 22 18:25:56 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	170	-0.02
2	1,4-Dioxane	0.400	0.345	13.8	160	-0.01
3	n-Nitrosodimethylamine	0.400	0.489	-22.2	196	-0.01
4 S	2-Fluorophenol	0.400	0.400	0.0	156	0.00
5 S	Phenol-d6	0.400	0.357	10.8	163	0.00
6	bis(2-Chloroethyl)ether	0.400	0.415	-3.7	195	-0.02
7 I	Naphthalene-d8	5.000	5.000	0.0	164	0.00
8 S	Nitrobenzene-d5	0.400	0.442	-10.5	178	0.00
9	Naphthalene	0.400	0.397	0.8	163	-0.02
10	Hexachlorobutadiene	0.400	0.384	4.0	159	-0.02
11	2-Methylnaphthalene	0.400	0.389	2.8	162	-0.01
12 I	Acenaphthene-d10	5.000	5.000	0.0	153	0.00
13 S	2,4,6-Tribromophenol	0.400	0.396	1.0	132	0.00
14 S	2-Fluorobiphenyl	0.400	0.423	-5.7	164	0.00
15	Acenaphthylene	0.400	0.391	2.3	153	0.00
16	Acenaphthene	0.400	0.408	-2.0	151	0.00
17	Fluorene	0.400	0.389	2.8	151	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	136	0.02
19	Hexachlorobenzene	0.400	0.439	-9.7	159	0.00
20	Pentachlorophenol	0.400	0.401	-0.3	145	0.00
21	Phenanthrene	0.400	0.460	-15.0	165	0.00
22	Anthracene	0.400	0.437	-9.2	163	0.02
23	Fluoranthene	0.400	0.402	-0.5	148	0.00
24 I	Chrysene-d12	5.000	5.000	0.0	142	0.00
25	Pyrene	0.400	0.446	-11.5	165	0.00
26 S	Terphenyl-d14	0.400	0.409	-2.2	147	0.02
27	Benzo(a)anthracene	0.400	0.424	-6.0	154	0.00
28	Chrysene	0.400	0.443	-10.7	147	0.00
29	Bis(2-ethylhexyl)phthalate	0.400	0.380	5.0	133	0.00
30	Indeno(1,2,3-cd)pyrene	0.400	0.368	8.0	133	0.04
31 I	Perylene-d12	5.000	5.000	0.0	136	0.02
32	Benzo(b)fluoranthene	0.400	0.408	-2.0	139	0.02
33	Benzo(k)fluoranthene	0.400	0.402	-0.5	147	0.02
34 C	Benzo(a)pyrene	0.400	0.398	0.5	143	0.02
35	Dibenzo(a,h)anthracene	0.400	0.381	4.8	135	0.04
36	Benzo(g,h,i)perylene	0.400	0.379	5.3	133	0.04

(#) = Out of Range

SPCC's out = 0 CCC's out = 0