

Data Path : Z:\HPCHEM1\BNA_E\Data\Be102116\
 Data File : BE092469.D
 Acq On : 21 Oct 2016 16:23
 Operator : UM/SJ
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4EC

Quant Time: Oct 21 17:30:58 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE092216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 21 17:26:45 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	159	-0.02
2	1,4-Dioxane	0.400	0.355	11.3	153	-0.01
3	n-Nitrosodimethylamine	0.400	0.478	-19.5	177	-0.01
4 S	2-Fluorophenol	0.400	0.403	-0.8	148	0.00
5 S	Phenol-d6	0.400	0.360	10.0	154	0.00
6	bis(2-Chloroethyl)ether	0.400	0.412	-3.0	181	-0.02
7 I	Naphthalene-d8	5.000	5.000	0.0	158	0.00
8 S	Nitrobenzene-d5	0.400	0.443	-10.7	172	0.00
9	Naphthalene	0.400	0.396	1.0	157	-0.02
10	Hexachlorobutadiene	0.400	0.384	4.0	153	-0.02
11	2-Methylnaphthalene	0.400	0.392	2.0	157	-0.01
12 I	Acenaphthene-d10	5.000	5.000	0.0	148	0.00
13 S	2,4,6-Tribromophenol	0.400	0.404	-1.0	132	0.00
14 S	2-Fluorobiphenyl	0.400	0.424	-6.0	159	0.00
15	Acenaphthylene	0.400	0.392	2.0	149	0.00
16	Acenaphthene	0.400	0.410	-2.5	147	0.00
17	Fluorene	0.400	0.390	2.5	146	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	133	0.02
19	Hexachlorobenzene	0.400	0.431	-7.7	153	0.00
20	Pentachlorophenol	0.400	0.387	3.3	132	0.00
21	Phenanthrene	0.400	0.449	-12.2	157	0.00
22	Anthracene	0.400	0.447	-11.7	163	0.02
23	Fluoranthene	0.400	0.403	-0.8	145	0.00
24 I	Chrysene-d12	5.000	5.000	0.0	146	0.00
25	Pyrene	0.400	0.429	-7.2	162	0.00
26 S	Terphenyl-d14	0.400	0.408	-2.0	150	0.00
27	Benzo(a)anthracene	0.400	0.400	0.0	149	0.00
28	Chrysene	0.400	0.409	-2.2	139	0.00
29	Bis(2-ethylhexyl)phthalate	0.400	0.345	13.8	124	0.00
30	Indeno(1,2,3-cd)pyrene	0.400	0.364	9.0	135	0.03
31 I	Perylene-d12	5.000	5.000	0.0	134	0.03
32	Benzo(b)fluoranthene	0.400	0.394	1.5	133	0.02
33	Benzo(k)fluoranthene	0.400	0.420	-5.0	151	0.03
34 C	Benzo(a)pyrene	0.400	0.401	-0.3	142	0.02
35	Dibenzo(a,h)anthracene	0.400	0.391	2.3	137	0.05
36	Benzo(g,h,i)perylene	0.400	0.379	5.3	131	0.05

(#) = Out of Range

SPCC's out = 0 CCC's out = 0