

Data Path : Z:\HPCHEM1\BNA_E\Data\Be102116\
 Data File : BE092469.D
 Acq On : 21 Oct 2016 16:23
 Operator : UM/SJ
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4EC

Quant Time: Oct 21 17:30:58 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE092216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 21 17:26:45 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	159#	-0.02
2	1,4-Dioxane	0.635	0.617	2.8	153#	-0.01
3	n-Nitrosodimethylamine	0.588	0.591	-0.5	177#	-0.01
4 S	2-Fluorophenol	0.999	0.867	13.2	148	0.00
5 S	Phenol-d6	1.388	1.170	15.7	154#	0.00
6	bis(2-Chloroethyl)ether	1.222	1.084	11.3	181#	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	158#	0.00
8 S	Nitrobenzene-d5	0.299	0.276	7.7	172#	0.00
9	Naphthalene	1.082	1.072	0.9	157#	-0.02
10	Hexachlorobutadiene	0.170	0.164	3.5	153#	-0.02
11	2-Methylnaphthalene	0.716	0.702	2.0	157#	-0.01
12 I	Acenaphthene-d10	1.000	1.000	0.0	148	0.00
13 S	2,4,6-Tribromophenol	0.144	0.110	23.6	132	0.00
14 S	2-Fluorobiphenyl	1.184	1.256	-6.1	159#	0.00
15	Acenaphthylene	7.240	7.099	1.9	149	0.00
16	Acenaphthene	1.134	1.162	-2.5	147	0.00
17	Fluorene	1.438	1.403	2.4	146	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	133	0.02
19	Hexachlorobenzene	0.214	0.231	-7.9	153#	0.00
20	Pentachlorophenol	0.050	0.042	16.0	132	0.00
21	Phenanthrene	1.149	1.289	-12.2	157#	0.00
22	Anthracene	1.105	1.236	-11.9	163#	0.02
23	Fluoranthene	5.460	5.498	-0.7	145	0.00
24 I	Chrysene-d12	1.000	1.000	0.0	146	0.00
25	Pyrene	4.333	4.643	-7.2	162#	0.00
26 S	Terphenyl-d14	0.620	0.632	-1.9	150#	0.00
27	Benzo(a)anthracene	4.873	4.867	0.1	149	0.00
28	Chrysene	4.462	4.563	-2.3	139	0.00
29	Bis(2-ethylhexyl)phthalate	0.420	0.362	13.8	124	0.00
30	Indeno(1,2,3-cd)pyrene	4.734	4.310	9.0	135	0.03
31 I	Perylene-d12	1.000	1.000	0.0	134	0.03
32	Benzo(b)fluoranthene	1.211	1.194	1.4	133	0.02
33	Benzo(k)fluoranthene	1.267	1.330	-5.0	151#	0.03
34 C	Benzo(a)pyrene	1.187	1.189	-0.2	142	0.02
35	Dibenzo(a,h)anthracene	1.088	1.064	2.2	137	0.05
36	Benzo(g,h,i)perylene	4.673	4.426	5.3	131	0.05

(#) = Out of Range

SPCC's out = 0 CCC's out = 0