

Data Path : Z:\HPCHEM1\BNA_E\Data\BE102214\
 Data File : BE088180.D
 Acq On : 22 Oct 2014 13:40
 Operator : TP/JJ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Oct 27 04:18:20 2014
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-PAH-BE102014.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 21 02:11:14 2014
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)

1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	77	0.00
2	1,4-Dioxane	1.000	1.011	-1.1	74	-0.06
3 I	Naphthalene-d8	5.000	5.000	0.0	78	0.00
4 S	Nitrobenzene-d5	1.000	1.027	-2.7	77	0.00
5	Naphthalene	1.000	1.036	-3.6	77	0.00
6	2-Methylnaphthalene	1.000	1.048	-4.8	78	0.00
7 I	Acenaphthene-d10	5.000	5.000	0.0	78	0.00
8 S	2-Fluorobiphenyl	1.000	1.055	-5.5	79	0.00
9	Acenaphthylene	1.000	1.062	-6.2	80	0.01
10	Acenaphthene	1.000	1.022	-2.2	79	0.00
11	Fluorene	1.000	1.019	-1.9	77	0.00
12 I	Phenanthrene-d10	5.000	5.000	0.0	73	0.00
13	Phenanthrene	1.000	0.926	7.4	59	0.00
14	Anthracene	1.000	0.996	0.4	71	0.00
15	Fluoranthene	1.000	0.862	13.8	63	0.00
16 I	Chrysene-d12	5.000	5.000	0.0	60	0.00
17	Pyrene	1.000	1.065	-6.5	64	0.00
18 S	Terphenyl-d14	1.000	1.080	-8.0	64	0.00
19	Benzo(a)anthracene	1.000	0.946	5.4	58	0.00
20	Chrysene	1.000	1.018	-1.8	61	0.00
21	Indeno(1,2,3-cd)pyrene	1.000	1.025	-2.5	62	0.00
22 I	Perylene-d12	5.000	5.000	0.0	59	0.00
23	Benzo(b)fluoranthene	1.000	0.914	8.6	54	0.00
24	Benzo(k)fluoranthene	1.000	1.003	-0.3	59	0.00
25 C	Benzo(a)pyrene	1.000	0.959	4.1	55	0.00
26	Dibenzo(a,h)anthracene	1.000	0.988	1.2	58	0.00
27	Benzo(g,h,i)perylene	1.000	1.042	-4.2	62	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0