

Data Path : Z:\HPCHEM1\BNA_E\Data\BE102214\
 Data File : BE088180.D
 Acq On : 22 Oct 2014 13:40
 Operator : TP/JJ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Oct 27 04:18:20 2014
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-PAH-BE102014.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 21 02:11:14 2014
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)

1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	77	0.00
2	1,4-Dioxane	0.719	0.728	-1.3	74	-0.06
3 I	Naphthalene-d8	1.000	1.000	0.0	78	0.00
4 S	Nitrobenzene-d5	0.360	0.370	-2.8	77	0.00
5	Naphthalene	1.031	1.068	-3.6	77	0.00
6	2-Methylnaphthalene	0.617	0.647	-4.9	78	0.00
7 I	Acenaphthene-d10	1.000	1.000	0.0	78	0.00
8 S	2-Fluorobiphenyl	1.381	1.457	-5.5	79	0.00
9	Acenaphthylene	7.619	8.092	-6.2	80	0.01
10	Acenaphthene	1.176	1.201	-2.1	79	0.00
11	Fluorene	1.246	1.270	-1.9	77	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	73	0.00
13	Phenanthrene	4.779	4.089	14.4	59	0.00
14	Anthracene	3.902	3.885	0.4	71	0.00
15	Fluoranthene	0.951	0.820	13.8	63	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	60	0.00
17	Pyrene	1.359	1.447	-6.5	64	0.00
18 S	Terphenyl-d14	0.805	0.869	-8.0	64	0.00
19	Benzo(a)anthracene	4.447	4.208	5.4	58	0.00
20	Chrysene	4.363	4.443	-1.8	61	0.00
21	Indeno(1,2,3-cd)pyrene	4.846	4.965	-2.5	62	0.00
22 I	Perylene-d12	1.000	1.000	0.0	59	0.00
23	Benzo(b)fluoranthene	1.190	1.087	8.7	54	0.00
24	Benzo(k)fluoranthene	1.274	1.278	-0.3	59	0.00
25 C	Benzo(a)pyrene	1.096	1.051	4.1	55	0.00
26	Dibenzo(a,h)anthracene	1.075	1.062	1.2	58	0.00
27	Benzo(g,h,i)perylene	4.752	4.954	-4.3	62	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0