

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE102315\
 Data File : BE091041.D
 Acq On : 23 Oct 2015 19:45
 Operator : UM/NP
 Sample : G4143-08
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 YT-MW-01

Quant Time: Oct 24 04:58:15 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE092815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Oct 24 02:19:56 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.91	152	20127	5.00	ng	0.00
7) Naphthalene-d8	9.58	136	104914	5.00	ng	-0.01
13) Acenaphthene-d10	13.41	164	64375	5.00	ng	0.00
19) Phenanthrene-d10	16.13	188	138158	5.00	ng	0.00
26) Chrysene-d12	20.29	240	160930	5.00	ng	0.00
35) Perylene-d12	22.19	264	158386	5.00	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	4.92	112	18140	4.48	ng	-0.01
5) Phenol-d6	6.55	99	14303	2.78	ng	0.00
8) Nitrobenzene-d5	8.17	82	43779	6.28	ng	-0.01
14) 2,4,6-Tribromophenol	14.94	330	20551	9.19	ng	0.00
15) 2-Fluorobiphenyl	12.03	172	79564	4.56	ng	-0.01
29) Terphenyl-d14	18.69	244	112249	5.65	ng	0.00
Target Compounds						
10) Naphthalene	9.62	128	39303	1.90	ng	# 92
12) 2-Methylnaphthalene	11.15	142	105629	7.93	ng	97
18) Fluorene	14.46	166	64840	3.05	ng	# 66
23) Phenanthrene	16.17	178	107209	3.39	ng	# 73
25) Fluoranthene	18.16	202	20468	0.54	ng	# 91
28) Pyrene	18.54	202	32057	0.86	ng	# 80
30) Benzo(a)anthracene	20.26	228	6676	0.19	ng	# 83
33) Bis(2-ethylhexyl)phthalate	19.97	149	11893	1.01	ng	99
34) Indeno(1,2,3-cd)pyrene	23.84	276	2726	0.18	ng	93
36) Benzo(b)fluoranthene	21.63	252	4351	0.12	ng	# 52
38) Benzo(a)pyrene	22.09	252	4719	0.15	ng	# 54
40) Benzo(g,h,i)perylene	24.40	276	2645	0.08	ng	# 62

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE102315\
Data File : BE091041.D
Acq On : 23 Oct 2015 19:45
Operator : UM/NP
Sample : G4143-08
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
YT-MW-01

Quant Time: Oct 24 04:58:15 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE092815.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Oct 24 02:19:56 2015
Response via : Initial Calibration

