

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE102315\
 Data File : BE091035.D
 Acq On : 23 Oct 2015 15:55
 Operator : UM/NP
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Oct 26 15:44:51 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE092815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 07 17:11:43 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	51	-0.04
2	1,4-Dioxane	0.888	0.621	30.1#	42#	-0.13
3	n-Nitrosodimethylamine	0.879	1.033	-17.5	66	-0.21
4 S	2-Fluorophenol	1.045	1.110	-6.2	62	-0.06
5 S	Phenol-d6	1.280	1.370	-7.0	61	-0.07
6	bis(2-Chloroethyl)ether	1.524	2.558	-67.8#	89	-0.09
7 I	Naphthalene-d8	1.000	1.000	0.0	51	-0.05
8 S	Nitrobenzene-d5	0.270	0.399	-47.8#	79	-0.06
9	Nitrobenzene	0.289	0.424	-46.7#	83	-0.06
10	Naphthalene	1.075	1.033	3.9	51	-0.05
11	Hexachlorobutadiene	0.179	0.149	16.8	42#	-0.04
12	2-Methylnaphthalene	0.634	0.661	-4.3	56	-0.05
13 I	Acenaphthene-d10	1.000	1.000	0.0	48#	-0.04
14 S	2,4,6-Tribromophenol	0.158	0.135	14.6	55	-0.04
15 S	2-Fluorobiphenyl	1.355	1.336	1.4	51	-0.05
16	Acenaphthylene	8.351	8.161	2.3	52	-0.05
17	Acenaphthene	1.334	1.273	4.6	49#	-0.04
18	Fluorene	1.649	1.648	0.1	52	-0.05
19 I	Phenanthrene-d10	1.000	1.000	0.0	51	-0.05
20	4-Bromophenyl-phenylether	0.189	0.168	11.1	49#	-0.05
21	Hexachlorobenzene	0.970	0.199	79.5#	11#	-0.05
22	Pentachlorophenol	0.070	0.047	32.9#	53	-0.04
23	Phenanthrene	1.265	1.187	6.2	52	-0.05
24	Anthracene	1.131	1.096	3.1	56	-0.05
25	Fluoranthene	1.503	1.366	9.1	57	-0.05
26 I	Chrysene-d12	1.000	1.000	0.0	54	-0.04
27	Benzydine	0.090	0.161	-78.9#	119	-0.07
28	Pyrene	1.277	1.097	14.1	56	-0.05
29 S	Terphenyl-d14	0.631	0.674	-6.8	66	-0.05
30	Benzo(a)anthracene	1.176	1.123	4.5	57	-0.04
31	3,3'-Dichlorobenzidine	0.201	0.000	100.0#	0#	-20.84#
32	Chrysene	1.505	1.214	19.3	51	-0.04
33	Bis(2-ethylhexyl)phthalate	0.469	0.402	14.3	73	-0.04
34	Indeno(1,2,3-cd)pyrene	1.212	1.205	0.6	62	-0.09
35 I	Perylene-d12	1.000	1.000	0.0	61	-0.06
36	Benzo(b)fluoranthene	1.102	1.136	-3.1	70	-0.05
37	Benzo(k)fluoranthene	1.515	1.215	19.8	55	-0.05
38 C	Benzo(a)pyrene	1.022	1.062	-3.9	74	-0.05
39	Dibenzo(a,h)anthracene	1.009	1.138	-12.8	82	-0.09
40	Benzo(g,h,i)perylene	1.096	1.231	-12.3	79	-0.09

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE102315\
Data File : BE091035.D
Acq On : 23 Oct 2015 15:55
Operator : UM/NP
Sample : SSTDCCC001
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC001

Quant Time: Oct 26 15:44:51 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE092815.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 07 17:11:43 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				