

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE102315\
 Data File : BE091035.D
 Acq On : 23 Oct 2015 15:55
 Operator : UM/NP
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Oct 26 15:44:51 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE092815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 07 17:11:43 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	51	-0.04
2	1,4-Dioxane	1.000	0.759	24.1	42	-0.13
3	n-Nitrosodimethylamine	1.000	1.290	-29.0#	66	-0.21
4 S	2-Fluorophenol	1.000	1.156	-15.6	62	-0.06
5 S	Phenol-d6	1.000	1.070	-7.0	61	-0.07
6	bis(2-Chloroethyl)ether	1.000	1.905	-90.5#	89	-0.09
7 I	Naphthalene-d8	5.000	5.000	0.0	51	-0.05
8 S	Nitrobenzene-d5	1.000	1.347	-34.7#	79	-0.06
9	Nitrobenzene	1.000	1.346	-34.6#	83	-0.06
10	Naphthalene	1.000	1.008	-0.8	51	-0.05
11	Hexachlorobutadiene	1.000	0.842	15.8	42	-0.04
12	2-Methylnaphthalene	1.000	1.041	-4.1	56	-0.05
13 I	Acenaphthene-d10	5.000	5.000	0.0	48	-0.04
14 S	2,4,6-Tribromophenol	1.000	0.932	6.8	55	-0.04
15 S	2-Fluorobiphenyl	1.000	0.986	1.4	51	-0.05
16	Acenaphthylene	1.000	0.977	2.3	52	-0.05
17	Acenaphthene	1.000	0.954	4.6	49	-0.04
18	Fluorene	1.000	0.999	0.1	52	-0.05
19 I	Phenanthrene-d10	5.000	5.000	0.0	51	-0.05
20	4-Bromophenyl-phenylether	1.000	0.887	11.3	49	-0.05
21	Hexachlorobenzene	1.000	0.119	88.1#	11	-0.05
22	Pentachlorophenol	1.000	0.865	13.5	53	-0.04
23	Phenanthrene	1.000	0.988	1.2	52	-0.05
24	Anthracene	1.000	1.017	-1.7	56	-0.05
25	Fluoranthene	1.000	1.023	-2.3	57	-0.05
26 I	Chrysene-d12	5.000	5.000	0.0	54	-0.04
27	Benzydine	1.000	1.891	-89.1#	119	-0.07
28	Pyrene	1.000	0.945	5.5	56	-0.05
29 S	Terphenyl-d14	1.000	1.117	-11.7	66	-0.05
30	Benzo(a)anthracene	1.000	1.011	-1.1	57	-0.04
31	3,3'-Dichlorobenzidine	1.000	0.000	100.0#	0	-20.84#
32	Chrysene	1.000	0.908	9.2	51	-0.04
33	Bis(2-ethylhexyl)phthalate	1.000	1.088	-8.8	73	-0.04
34	Indeno(1,2,3-cd)pyrene	1.000	1.051	-5.1	62	-0.09
35 I	Perylene-d12	5.000	5.000	0.0	61	-0.06
36	Benzo(b)fluoranthene	1.000	1.030	-3.0	70	-0.05
37	Benzo(k)fluoranthene	1.000	0.837	16.3	55	-0.05
38 C	Benzo(a)pyrene	1.000	1.039	-3.9	74	-0.05
39	Dibenzo(a,h)anthracene	1.000	1.127	-12.7	82	-0.09
40	Benzo(g,h,i)perylene	1.000	1.123	-12.3	79	-0.09

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE102315\
Data File : BE091035.D
Acq On : 23 Oct 2015 15:55
Operator : UM/NP
Sample : SSTDCCC001
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC001

Quant Time: Oct 26 15:44:51 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE092815.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 07 17:11:43 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev Area	% Dev(min)
----------	--------	-------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0