

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE102315\
 Data File : BE091045.D
 Acq On : 23 Oct 2015 22:08
 Operator : UM/NP
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Oct 24 03:33:16 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE092815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Oct 24 03:31:05 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	47#	-0.05
2	1,4-Dioxane	0.888	0.690	22.3	42#	-0.18
3	n-Nitrosodimethylamine	0.879	0.780	11.3	46#	-0.27
4 S	2-Fluorophenol	1.045	1.325	-26.8#	67	-0.07
5 S	Phenol-d6	1.280	1.746	-36.4#	71	-0.07
6	bis(2-Chloroethyl)ether	1.524	2.422	-58.9#	77	-0.11
7 I	Naphthalene-d8	1.000	1.000	0.0	48#	-0.05
8 S	Nitrobenzene-d5	0.270	0.422	-56.3#	79	-0.06
9	Nitrobenzene	0.289	0.445	-54.0#	82	-0.06
10	Naphthalene	1.075	1.037	3.5	48#	-0.06
11	Hexachlorobutadiene	0.179	0.149	16.8	40#	-0.05
12	2-Methylnaphthalene	0.634	0.695	-9.6	55	-0.06
13 I	Acenaphthene-d10	1.000	1.000	0.0	51	-0.05
14 S	2,4,6-Tribromophenol	0.158	0.202	-27.8#	87	-0.05
15 S	2-Fluorobiphenyl	1.355	1.279	5.6	52	-0.06
16	Acenaphthylene	8.351	8.529	-2.1	57	-0.05
17	Acenaphthene	1.334	1.288	3.4	53	-0.05
18	Fluorene	1.649	1.685	-2.2	56	-0.05
19 I	Phenanthrene-d10	1.000	1.000	0.0	57	-0.05
20	4-Bromophenyl-phenylether	0.189	0.173	8.5	57	-0.05
21	Hexachlorobenzene	0.970	0.188	80.6#	12#	-0.05
22	Pentachlorophenol	0.070	0.083	-18.6	106	-0.04
23	Phenanthrene	1.265	1.198	5.3	59	-0.05
24	Anthracene	1.131	1.138	-0.6	65	-0.05
25	Fluoranthene	1.503	1.452	3.4	68	-0.05
26 I	Chrysene-d12	1.000	1.000	0.0	60	-0.05
27	Benzydine	0.090	0.208	-131.1#	170#	-0.08
28	Pyrene	1.277	1.195	6.4	67	-0.05
29 S	Terphenyl-d14	0.631	0.726	-15.1	79	-0.05
30	Benzo(a)anthracene	1.176	1.213	-3.1	69	-0.05
31	3,3'-Dichlorobenzidine	0.201	0.000	100.0#	0#	-20.84#
32	Chrysene	1.505	1.231	18.2	58	-0.04
33	Bis(2-ethylhexyl)phthalate	0.469	0.578	-23.2	116	-0.05
34	Indeno(1,2,3-cd)pyrene	1.212	1.209	0.2	69	-0.08
35 I	Perylene-d12	1.000	1.000	0.0	66	-0.06
36	Benzo(b)fluoranthene	1.102	1.222	-10.9	82	-0.05
37	Benzo(k)fluoranthene	1.515	1.218	19.6	59	-0.05
38 C	Benzo(a)pyrene	1.022	1.126	-10.2	85	-0.05
39	Dibenzo(a,h)anthracene	1.009	1.190	-17.9	93	-0.08
40	Benzo(g,h,i)perylene	1.096	1.243	-13.4	87	-0.09

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE102315\
Data File : BE091045.D
Acq On : 23 Oct 2015 22:08
Operator : UM/NP
Sample : SSTDCCC001
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC001

Quant Time: Oct 24 03:33:16 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE092815.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Oct 24 03:31:05 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				