

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE102315\
 Data File : BE091045.D
 Acq On : 23 Oct 2015 22:08
 Operator : UM/NP
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Oct 24 03:33:16 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE092815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Oct 24 03:31:05 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	47	-0.05
2	1,4-Dioxane	1.000	0.872	12.8	42	-0.18
3	n-Nitrosodimethylamine	1.000	0.963	3.7	46	-0.27
4 S	2-Fluorophenol	1.000	1.367	-36.7#	67	-0.07
5 S	Phenol-d6	1.000	1.364	-36.4#	71	-0.07
6	bis(2-Chloroethyl)ether	1.000	1.806	-80.6#	77	-0.11
7 I	Naphthalene-d8	5.000	5.000	0.0	48	-0.05
8 S	Nitrobenzene-d5	1.000	1.413	-41.3#	79	-0.06
9	Nitrobenzene	1.000	1.400	-40.0#	82	-0.06
10	Naphthalene	1.000	1.012	-1.2	48	-0.06
11	Hexachlorobutadiene	1.000	0.842	15.8	40	-0.05
12	2-Methylnaphthalene	1.000	1.095	-9.5	55	-0.06
13 I	Acenaphthene-d10	5.000	5.000	0.0	51	-0.05
14 S	2,4,6-Tribromophenol	1.000	1.314	-31.4#	87	-0.05
15 S	2-Fluorobiphenyl	1.000	0.944	5.6	52	-0.06
16	Acenaphthylene	1.000	1.021	-2.1	57	-0.05
17	Acenaphthene	1.000	0.965	3.5	53	-0.05
18	Fluorene	1.000	1.022	-2.2	56	-0.05
19 I	Phenanthrene-d10	5.000	5.000	0.0	57	-0.05
20	4-Bromophenyl-phenylether	1.000	0.917	8.3	57	-0.05
21	Hexachlorobenzene	1.000	0.106	89.4#	12	-0.05
22	Pentachlorophenol	1.000	1.474	-47.4#	106	-0.04
23	Phenanthrene	1.000	0.998	0.2	59	-0.05
24	Anthracene	1.000	1.057	-5.7	65	-0.05
25	Fluoranthene	1.000	1.089	-8.9	68	-0.05
26 I	Chrysene-d12	5.000	5.000	0.0	60	-0.05
27	Benzydine	1.000	2.260	-126.0#	170	-0.08
28	Pyrene	1.000	1.031	-3.1	67	-0.05
29 S	Terphenyl-d14	1.000	1.201	-20.1	79	-0.05
30	Benzo(a)anthracene	1.000	1.091	-9.1	69	-0.05
31	3,3'-Dichlorobenzidine	1.000	0.022	97.8#	0	-20.84#
32	Chrysene	1.000	0.923	7.7	58	-0.04
33	Bis(2-ethylhexyl)phthalate	1.000	1.497	-49.7#	116	-0.05
34	Indeno(1,2,3-cd)pyrene	1.000	1.055	-5.5	69	-0.08
35 I	Perylene-d12	5.000	5.000	0.0	66	-0.06
36	Benzo(b)fluoranthene	1.000	1.109	-10.9	82	-0.05
37	Benzo(k)fluoranthene	1.000	0.839	16.1	59	-0.05
38 C	Benzo(a)pyrene	1.000	1.102	-10.2	85	-0.05
39	Dibenzo(a,h)anthracene	1.000	1.179	-17.9	93	-0.08
40	Benzo(g,h,i)perylene	1.000	1.134	-13.4	87	-0.09

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE102315\
Data File : BE091045.D
Acq On : 23 Oct 2015 22:08
Operator : UM/NP
Sample : SSTDCCC001
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC001

Quant Time: Oct 24 03:33:16 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE092815.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Oct 24 03:31:05 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				