

Data Path : Z:\HPCHEM1\BNA E\DATA\BE102317\
 Data File : BE094509.D
 Acq On : 23 Oct 2017 23:26
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 24 02:22:44 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE102317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 23 22:40:47 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	0.00
2	1,4-Dioxane	0.677	0.753	-11.2	95	-0.01
3	n-Nitrosodimethylamine	0.648	0.666	-2.8	101	-0.01
4 S	2-Fluorophenol	1.370	1.457	-6.4	105	-0.01
5 S	Phenol-d6	2.069	2.400	-16.0	117	-0.01
6	bis(2-Chloroethyl)ether	1.600	1.607	-0.4	100	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	101	-0.02
8 S	Nitrobenzene-d5	0.319	0.323	-1.3	106	0.00
9	Naphthalene	4.527	4.453	1.6	97	0.00
10	Hexachlorobutadiene	0.164	0.164	0.0	100	0.00
11 SURR	2-Methylnaphthalene-d10	0.642	0.634	1.2	97	0.00
12	2-Methylnaphthalene	0.768	0.768	0.0	100	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	99	0.00
14 S	2,4,6-Tribromophenol	0.110	0.136	-23.6	127	0.00
15 S	2-Fluorobiphenyl	1.410	1.411	-0.1	98	0.00
16	Acenaphthylene	2.112	2.109	0.1	100	0.00
17	Acenaphthene	1.356	1.348	0.6	97	-0.01
18	Fluorene	1.738	1.700	2.2	97	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	112	0.00
20	4-Bromophenyl-phenylether	0.218	0.225	-3.2	117	0.00
21	Hexachlorobenzene	0.191	0.180	5.8	91	0.00
22	Pentachlorophenol	0.024	0.028	-16.7	128	-0.03
23	Phenanthrene	1.141	0.930	18.5	90	0.00
24	Anthracene	1.176	1.060	9.9	98	-0.03
25 SURR	Fluoranthene-d10	4.676	4.886	-4.5	102	0.00
26	Fluoranthene	1.420	1.469	-3.5	101	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	99	-0.01
28	Pvrene	1.399	1.404	-0.4	101	0.00
29 S	Terphenyl-d14	0.754	0.753	0.1	101	0.00
30	Benzo(a)anthracene	1.302	1.302	0.0	99	0.00
31	Chrysene	1.322	1.335	-1.0	103	0.00
32	Bis(2-ethylhexyl)phthalate	3.193	3.191	0.1	108	-0.01
33	Indeno(1,2,3-cd)pyrene	1.222	1.265	-3.5	104	0.00
34 I	Perylene-d12	1.000	1.000	0.0	100	0.00
35	Benzo(b)fluoranthene	1.294	1.279	1.2	101	0.00
36	Benzo(k)fluoranthene	1.379	1.355	1.7	99	0.00
37 C	Benzo(a)pyrene	1.255	1.243	1.0	99	0.00
38	Dibenzo(a,h)anthracene	1.127	1.140	-1.2	103	0.00
39	Benzo(a,h,i)perylene	1.055	1.096	-3.9	105	0.00

Data Path : Z:\HPCHEM1\BNA E\DATA\BE102317\
Data File : BE094509.D
Acq On : 23 Oct 2017 23:26
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 24 02:22:44 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE102317.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 23 22:40:47 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			