

Data Path : Z:\HPCHEM1\BNA E\DATA\BE102417\
 Data File : BE094513.D
 Acq On : 24 Oct 2017 11:35
 Operator : SJ/JU
 Sample : I5950-10DL 10X
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 20170723-C-A-COMP2DL

Manual Integrations
 APPROVED

Sohil
 10/25/2017 6:24:05 PM

Quant Time: Oct 24 16:06:21 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE102317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 24 15:59:47 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.91	152	1237	0.40	ng	0.00
7) Naphthalene-d8	10.69	136	6826	0.40	ng	-0.02
13) Acenaphthene-d10	14.51	164	4273	0.40	ng	0.00
19) Phenanthrene-d10	17.24	188	12751	0.40	ng	0.00
27) Chrysene-d12	21.41	240	10166	0.40	ng	-0.01
34) Perylene-d12	23.95	264	9639	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.54	112	73	0.02	ng	0.02
5) Phenol-d6	7.17	99	64	0.01	ng	0.03
8) Nitrobenzene-d5	9.10	82	229	0.04	ng	0.02
11) 2-Methylnaphthalene-d10	12.28	152	460	0.04	ng	0.00
14) 2,4,6-Tribromophenol	16.03	330	57	0.05	ng	0.00
15) 2-Fluorobiphenyl	13.15	172	719	0.05	ng	0.00
25) Fluoranthene-d10	19.28	212	5443	0.04	ng	0.00
29) Terphenyl-d14	19.85	244	1027	0.05	ng	-0.01
Target Compounds						
9) Naphthalene	10.74	128	26326	0.341	ng	Qvalue 100
12) 2-Methylnaphthalene	12.36	142	2155	0.164	ng	99
17) Acenaphthene	14.57	154	1923	0.133	ng	99
18) Fluorene	15.56	166	2046	0.110	ng	# 92
23) Phenanthrene	17.27	178	7406m	0.186	ng	
24) Anthracene	17.37	178	1027m	0.027	ng	
28) Pyrene	19.67	202	1079	0.030	ng	# 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE102417\
Data File : BE094513.D
Acq On : 24 Oct 2017 11:35
Operator : SJ/JU
Sample : I5950-10DL 10X
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
20170723-C-A-COMP2DL

Manual Integrations
APPROVED

Sohil
10/25/2017 6:24:05 PM

Quant Time: Oct 24 16:06:21 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE102317.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 24 15:59:47 2017
Response via : Initial Calibration

