

Data Path : Z:\HPCHEM1\BNA E\DATA\BE102417\
 Data File : BE094514.D
 Acq On : 24 Oct 2017 12:11
 Operator : SJ/JU
 Sample : I5950-12DL 50X
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 20170725-PA-A-COMP1DL

Manual Integrations
 APPROVED

Sohil
 10/25/2017 6:24:06 PM

Quant Time: Oct 24 16:09:11 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE102317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 24 15:59:47 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.91	152	1121	0.40	ng	0.00
7) Naphthalene-d8	10.69	136	6493	0.40	ng	-0.02
13) Acenaphthene-d10	14.51	164	3931	0.40	ng	0.00
19) Phenanthrene-d10	17.24	188	8380	0.40	ng	0.00
27) Chrysene-d12	21.43	240	9189	0.40	ng	0.00
34) Perylene-d12	23.95	264	8787	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.54	112	11	0.00	ng	0.02
5) Phenol-d6	7.14	99	5	0.00	ng	0.00
8) Nitrobenzene-d5	9.10	82	43	0.01	ng	0.02
11) 2-Methylnaphthalene-d10	12.29	152	94	0.01	ng	0.00
14) 2,4,6-Tribromophenol	16.03	330	12	0.01	ng	0.00
15) 2-Fluorobiphenyl	13.16	172	132	0.01	ng	0.01
25) Fluoranthene-d10	19.29	212	947	0.01	ng	0.00
29) Terphenyl-d14	19.87	244	185	0.01	ng	0.00
Target Compounds						
9) Naphthalene	10.74	128	438765	5.971	ng	99
12) 2-Methylnaphthalene	12.36	142	11424	0.916	ng	98
17) Acenaphthene	14.58	154	7013	0.526	ng	99
18) Fluorene	15.58	166	5608	0.328	ng	97
23) Phenanthrene	17.30	178	9056m	0.407	ng	
24) Anthracene	17.37	178	1340m	0.054	ng	
28) Pyrene	19.67	202	700	0.022	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE102417\
Data File : BE094514.D
Acq On : 24 Oct 2017 12:11
Operator : SJ/JU
Sample : I5950-12DL 50X
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
20170725-PA-A-COMP1DL

Manual Integrations
APPROVED

Sohil
10/25/2017 6:24:06 PM

Quant Time: Oct 24 16:09:11 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE102317.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 24 15:59:47 2017
Response via : Initial Calibration

