

Data Path : Z:\HPCHEM1\BNA E\DATA\BE102417\
 Data File : BE094515.D
 Acq On : 24 Oct 2017 15:07
 Operator : SJ/JU
 Sample : I5950-12DL2 500X
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 20170725-PA-A-COMP1DL2

Manual Integrations
 APPROVED

Sohil
 10/25/2017 6:24:08 PM

Quant Time: Oct 24 16:58:27 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE102317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 23 22:40:47 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.92	152	1021	0.40	ng	0.01
7) Naphthalene-d8	10.71	136	5847	0.40	ng	0.00
13) Acenaphthene-d10	14.52	164	4329	0.40	ng	0.01
19) Phenanthrene-d10	17.27	188	9178	0.40	ng	0.03
27) Chrysene-d12	21.43	240	10343	0.40	ng	0.00
34) Perylene-d12	23.95	264	10041	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.51	112	2	0.00	ng	-0.01
5) Phenol-d6	7.16	99	2	0.00	ng	0.03
8) Nitrobenzene-d5	9.13	82	8	0.00	ng	0.05
11) 2-Methylnaphthalene-d10	12.31	152	4	0.00	ng	0.03
14) 2,4,6-Tribromophenol	16.06	330	2	0.00	ng	0.03
15) 2-Fluorobiphenyl	13.18	172	23	0.00	ng	0.03
25) Fluoranthene-d10	0.00	212	0d	0.00	ng	
29) Terphenyl-d14	19.87	244	54	0.00	ng	0.00
Target Compounds						
9) Naphthalene	10.76	128	41449m	0.626	ng	Qvalue
12) 2-Methylnaphthalene	12.41	142	1053	0.094	ng	92
17) Acenaphthene	14.58	154	736	0.050	ng	96
18) Fluorene	15.59	166	575	0.031	ng	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE102417\
Data File : BE094515.D
Acq On : 24 Oct 2017 15:07
Operator : SJ/JU
Sample : I5950-12DL2 500X
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
20170725-PA-A-COMP1DL2

Manual Integrations
APPROVED

Sohil
10/25/2017 6:24:08 PM

Quant Time: Oct 24 16:58:27 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE102317.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 23 22:40:47 2017
Response via : Initial Calibration

