


Data Path : Z:\HPCHEM1\BNA E\Data\BE102417\
 Data File : BE094510.D
 Acq On : 24 Oct 2017 9:22
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE102317.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Oct 23 22:40:47 2017



AutoFind: Scans 2565, 2566, 2567; Background Corrected with Scan 2555

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	41.4	8426	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	42.7	8687	PASS
70	69	0.00	2	0.4	38	PASS
127	198	10	80	48.4	9851	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	20349	PASS
199	198	5	9	6.3	1288	PASS
275	198	10	60	23.8	4837	PASS
365	198	1	100	2.6	525	PASS
441	442	0.01	24	16.3	2130	PASS
442	198	50	100	64.2	13055	PASS
443	442	15	24	19.5	2549	PASS