

Data Path : Z:\HPCHEM1\BNA E\DATA\BE102516\
 Data File : BE092479.D
 Acq On : 25 Oct 2016 14:43
 Operator : UM/SJ
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE102516

Quant Time: Oct 25 15:30:55 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE102516.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 25 15:25:33 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	89	0.00
2	1,4-Dioxane	0.654	0.625	4.4	90	0.00
3	n-Nitrosodimethylamine	0.600	0.532	11.3	82	0.00
4 S	2-Fluorophenol	0.758	0.801	-5.7	103	0.00
5 S	Phenol-d6	1.060	1.116	-5.3	100	0.00
6	bis(2-Chloroethyl)ether	1.251	1.179	5.8	89	0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	94	0.00
8 S	Nitrobenzene-d5	0.281	0.268	4.6	99	0.00
9	Naphthalene	1.081	1.031	4.6	89	0.00
10	Hexachlorobutadiene	0.166	0.162	2.4	91	0.00
11	2-Methylnaphthalene	0.698	0.666	4.6	93	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	95	0.00
13 S	2,4,6-Tribromophenol	0.098	0.095	3.1	100	0.00
14 S	2-Fluorobiphenyl	1.218	1.266	-3.9	101	0.00
15	Acenaphthylene	7.149	7.072	1.1	98	0.00
16	Acenaphthene	1.141	1.159	-1.6	98	0.00
17	Fluorene	1.366	1.377	-0.8	100	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	95	0.00
19	Hexachlorobenzene	0.224	0.229	-2.2	101	0.00
20	Pentachlorophenol	0.039	0.029	25.6#	83	0.00
21	Phenanthrene	1.292	1.238	4.2	100	0.00
22	Anthracene	1.163	1.074	7.7	100	0.00
23	Fluoranthene	5.122	4.999	2.4	99	0.00
24 I	Chrysene-d12	1.000	1.000	0.0	94	0.00
25	Pyrene	4.358	4.210	3.4	97	-0.02
26 S	Terphenyl-d14	0.619	0.614	0.8	101	0.00
27	Benzo(a)anthracene	4.626	4.314	6.7	93	0.00
28	Chrysene	4.862	5.093	-4.8	102	0.00
29	Bis(2-ethylhexyl)phthalate	0.385	0.320	16.9	103	0.00
30	Indeno(1,2,3-cd)pyrene	4.286	4.032	5.9	96	0.00
31 I	Perylene-d12	1.000	1.000	0.0	95	0.00
32	Benzo(b)fluoranthene	1.245	1.063	14.6	94	0.00
33	Benzo(k)fluoranthene	1.234	1.252	-1.5	94	0.00
34 C	Benzo(a)pyrene	1.127	0.978	13.2	94	0.00
35	Dibenzo(a,h)anthracene	1.053	0.968	8.1	96	0.00
36	Benzo(g,h,i)perylene	4.540	4.292	5.5	95	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0