

Data Path : Z:\HPCHEM1\BNA E\DATA\BE102516\
 Data File : BE092479.D
 Acq On : 25 Oct 2016 14:43
 Operator : UM/SJ
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE102516

Quant Time: Oct 25 15:30:55 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE102516.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 25 15:25:33 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	89	0.00
2	1,4-Dioxane	0.400	0.370	7.5	90	0.00
3	n-Nitrosodimethylamine	0.400	0.446	-11.5	82	0.00
4 S	2-Fluorophenol	0.400	0.425	-6.2	103	0.00
5 S	Phenol-d6	0.400	0.418	-4.5	100	0.00
6	bis(2-Chloroethyl)ether	0.400	0.430	-7.5	89	0.02
7 I	Naphthalene-d8	5.000	5.000	0.0	94	0.00
8 S	Nitrobenzene-d5	0.400	0.469	-17.2	99	0.00
9	Naphthalene	0.400	0.381	4.8	89	0.00
10	Hexachlorobutadiene	0.400	0.392	2.0	91	0.00
11	2-Methylnaphthalene	0.400	0.382	4.5	93	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	95	0.00
13 S	2,4,6-Tribromophenol	0.400	0.405	-1.3	100	0.00
14 S	2-Fluorobiphenyl	0.400	0.416	-4.0	101	0.00
15	Acenaphthylene	0.400	0.396	1.0	98	0.00
16	Acenaphthene	0.400	0.406	-1.5	98	0.00
17	Fluorene	0.400	0.403	-0.8	100	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	95	0.00
19	Hexachlorobenzene	0.400	0.410	-2.5	101	0.00
20	Pentachlorophenol	0.400	0.431	-7.7	83	0.00
21	Phenanthrene	0.400	0.383	4.3	100	0.00
22	Anthracene	0.400	0.369	7.8	100	0.00
23	Fluoranthene	0.400	0.390	2.5	99	0.00
24 I	Chrysene-d12	5.000	5.000	0.0	94	0.00
25	Pyrene	0.400	0.386	3.5	97	-0.02
26 S	Terphenyl-d14	0.400	0.397	0.8	101	0.00
27	Benzo(a)anthracene	0.400	0.373	6.8	93	0.00
28	Chrysene	0.400	0.419	-4.7	102	0.00
29	Bis(2-ethylhexyl)phthalate	0.400	0.448	-12.0	103	0.00
30	Indeno(1,2,3-cd)pyrene	0.400	0.376	6.0	96	0.00
31 I	Perylene-d12	5.000	5.000	0.0	95	0.00
32	Benzo(b)fluoranthene	0.400	0.341	14.8	94	0.00
33	Benzo(k)fluoranthene	0.400	0.406	-1.5	94	0.00
34 C	Benzo(a)pyrene	0.400	0.347	13.3	94	0.00
35	Dibenzo(a,h)anthracene	0.400	0.368	8.0	96	0.00
36	Benzo(g,h,i)perylene	0.400	0.378	5.5	95	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0