

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE102517\
 Data File : BE094523.D
 Acq On : 24 Oct 2017 20:20
 Operator : SJ/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SICV53

Quant Time: Oct 24 22:42:52 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE102517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 24 20:21:52 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	100	0.00
3	Naphthalene	0.978	0.983	-0.5	100	0.00
4 SURR	2-Methylnaphthalene-d10	0.649	0.656	-1.1	100	0.00
5	2-Methylnaphthalene	0.684	0.690	-0.9	100	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	98	0.00
7	Acenaphthylene	1.809	1.813	-0.2	98	0.00
8 C	Acenaphthene	1.300	1.305	-0.4	98	0.00
9	Fluorene	1.460	1.465	-0.3	98	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	96	0.00
11	Pentachlorophenol	0.020	0.016	20.0	95	0.02
12	Phenanthrene	1.057	1.079	-2.1	99	0.00
13	Anthracene	1.043	1.059	-1.5	99	0.00
14 SURR	Fluoranthene-d10	1.187	1.198	-0.9	98	0.00
15 C	Fluoranthene	1.221	1.233	-1.0	98	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	96	0.00
17	Pyrene	1.237	1.253	-1.3	98	0.00
18	Benzo(a)anthracene	1.143	1.166	-2.0	100	0.00
19	Chrysene	1.182	1.195	-1.1	97	0.00
20 I	Perylene-d12	1.000	1.000	0.0	95	0.00
21	Benzo(b)fluoranthene	1.125	1.144	-1.7	98	0.00
22	Benzo(k)fluoranthene	1.206	1.207	-0.1	99	0.00
23 C	Benzo(a)pyrene	1.125	1.138	-1.2	99	0.00
24	Indeno(1,2,3-cd)pyrene	1.164	1.160	0.3	98	0.00
25	Dibenzo(a,h)anthracene	0.981	0.978	0.3	97	0.00
26	Benzo(g,h,i)perylene	0.935	0.931	0.4	98	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0