

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE102517\
 Data File : BE094555.D
 Acq On : 26 Oct 2017 4:41
 Operator : SJ/JU
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.456

Quant Time: Oct 26 13:48:55 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE102517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 25 18:44:03 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	97	0.00
3	Naphthalene	0.978	0.942	3.7	94	0.00
4 SURR	2-Methylnaphthalene-d10	0.649	0.628	3.2	94	-0.01
5	2-Methylnaphthalene	0.684	0.658	3.8	93	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	90	0.00
7	Acenaphthylene	1.809	1.805	0.2	90	0.00
8 C	Acenaphthene	1.300	1.291	0.7	90	0.00
9	Fluorene	1.460	1.483	-1.6	92	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	91	0.00
11	Pentachlorophenol	0.020	0.022	-10.0	124	-0.05
12	Phenanthrene	1.057	1.033	2.3	90	-0.02
13	Anthracene	1.043	1.050	-0.7	93	0.00
14 SURR	Fluoranthene-d10	1.187	1.221	-2.9	95	0.00
15 C	Fluoranthene	1.221	1.253	-2.6	95	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	88	0.00
17	Pyrene	1.237	1.317	-6.5	95	0.00
18	Benzo(a)anthracene	1.143	1.200	-5.0	94	0.00
19	Chrysene	1.182	1.146	3.0	86	0.00
20 I	Perylene-d12	1.000	1.000	0.0	90	0.00
21	Benzo(b)fluoranthene	1.125	1.152	-2.4	93	0.00
22	Benzo(k)fluoranthene	1.206	1.156	4.1	89	0.00
23 C	Benzo(a)pyrene	1.125	1.154	-2.6	94	0.00
24	Indeno(1,2,3-cd)pyrene	1.164	1.229	-5.6	97	0.01
25	Dibenzo(a,h)anthracene	0.981	1.034	-5.4	97	0.01
26	Benzo(g,h,i)perylene	0.935	0.959	-2.6	95	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0