

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE102617\
 Data File : BE094584.D
 Acq On : 27 Oct 2017 4:24
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.459

Quant Time: Oct 27 10:12:26 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE102517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 27 04:05:19 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	99	0.00
2 I	Naphthalene-d8	0.400	0.400	0.0	99	0.00
3	Naphthalene	0.400	0.385	3.8	95	0.00
4 SURR	2-Methylnaphthalene-d10	0.400	0.388	3.0	95	0.00
5	2-Methylnaphthalene	0.400	0.381	4.8	93	0.00
6 I	Acenaphthene-d10	0.400	0.400	0.0	91	0.00
7	Acenaphthylene	0.400	0.419	-4.7	96	0.00
8 C	Acenaphthene	0.400	0.395	1.3	90	0.00
9	Fluorene	0.400	0.408	-2.0	93	0.00
10 I	Phenanthrene-d10	0.400	0.400	0.0	90	0.00
11	Pentachlorophenol	0.400	0.555	-38.8#	155	-0.03
12	Phenanthrene	0.400	0.405	-1.3	91	-0.02
13	Anthracene	0.400	0.410	-2.5	93	0.00
14 SURR	Fluoranthene-d10	0.400	0.406	-1.5	92	0.00
15 C	Fluoranthene	0.400	0.400	0.0	91	0.00
16 I	Chrysene-d12	0.400	0.400	0.0	85	0.00
17	Pyrene	0.400	0.427	-6.7	92	0.00
18	Benzo(a)anthracene	0.400	0.429	-7.2	93	0.00
19	Chrysene	0.400	0.391	2.3	83	0.00
20 I	Perylene-d12	0.400	0.400	0.0	87	0.00
21	Benzo(b)fluoranthene	0.400	0.415	-3.7	91	0.00
22	Benzo(k)fluoranthene	0.400	0.390	2.5	87	0.00
23 C	Benzo(a)pyrene	0.400	0.408	-2.0	91	0.00
24	Indeno(1,2,3-cd)pyrene	0.400	0.415	-3.7	93	0.00
25	Dibenzo(a,h)anthracene	0.400	0.425	-6.2	94	0.00
26	Benzo(g,h,i)perylene	0.400	0.408	-2.0	92	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0