

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE102714\
Data File : BE088326.D
Acq On : 27 Oct 2014 22:29
Operator : TP/JJ
Sample : PB79870BL
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
PB79870BL

Quant Time: Oct 28 02:06:52 2014
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-PAH-BE102714.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 28 01:58:22 2014
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.11	152	33552	5.00	ng	0.00
3) Naphthalene-d8	8.67	136	128672	5.00	ng	0.00
7) Acenaphthene-d10	10.81	164	61796	5.00	ng	0.00
12) Phenanthrene-d10	12.93	188	119103	5.00	ng	0.00
16) Chrysene-d12	18.34	240	67817	5.00	ng	0.00
22) Perylene-d12	21.40	264	63092	5.00	ng	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) Nitrobenzene-d5	7.79	82	76793	7.92	ng	0.00
8) 2-Fluorobiphenyl	10.01	172	131697	8.09	ng	0.00
18) Terphenyl-d14	16.10	244	102989	8.13	ng	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE102714\
Data File : BE088326.D
Acq On : 27 Oct 2014 22:29
Operator : TP/JJ
Sample : PB79870BL
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
PB79870BL

Quant Time: Oct 28 02:06:52 2014
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-PAH-BE102714.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 28 01:58:22 2014
Response via : Initial Calibration

