

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE102714\
 Data File : BE088325.D
 Acq On : 27 Oct 2014 21:56
 Operator : TP/JJ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE102714

Quant Time: Oct 28 10:23:36 2014
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-PAH-BE102714.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 28 10:16:06 2014
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)

1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	152	0.00
2	1,4-Dioxane	1.000	1.094	-9.4	162	0.09
3 I	Naphthalene-d8	5.000	5.000	0.0	145	0.00
4 S	Nitrobenzene-d5	1.000	1.087	-8.7	160	0.00
5	Naphthalene	1.000	1.081	-8.1	155	0.00
6	2-Methylnaphthalene	1.000	1.054	-5.4	149	0.00
7 I	Acenaphthene-d10	5.000	5.000	0.0	125	0.00
8 S	2-Fluorobiphenyl	1.000	1.128	-12.8	142	0.00
9	Acenaphthylene	1.000	1.153	-15.3	142	0.00
10	Acenaphthene	1.000	1.115	-11.5	137	-0.01
11	Fluorene	1.000	1.140	-14.0	137	0.00
12 I	Phenanthrene-d10	5.000	5.000	0.0	121	0.00
13	Phenanthrene	1.000	1.144	-14.4	140	0.00
14	Anthracene	1.000	1.103	-10.3	137	0.00
15	Fluoranthene	1.000	1.086	-8.6	137	0.00
16 I	Chrysene-d12	5.000	5.000	0.0	131	0.00
17	Pyrene	1.000	1.048	-4.8	139	0.00
18 S	Terphenyl-d14	1.000	1.051	-5.1	139	0.00
19	Benzo(a)anthracene	1.000	1.054	-5.4	150	0.00
20	Chrysene	1.000	1.059	-5.9	150	0.00
21	Indeno(1,2,3-cd)pyrene	1.000	1.061	-6.1	177	0.00
22 I	Perylene-d12	5.000	5.000	0.0	152	0.00
23	Benzo(b)fluoranthene	1.000	1.049	-4.9	168	0.00
24	Benzo(k)fluoranthene	1.000	1.092	-9.2	172	0.00
25 C	Benzo(a)pyrene	1.000	1.069	-6.9	174	0.00
26	Dibenzo(a,h)anthracene	1.000	1.044	-4.4	175	0.00
27	Benzo(g,h,i)perylene	1.000	1.045	-4.5	173	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0