

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE102714\
 Data File : BE088325.D
 Acq On : 27 Oct 2014 21:56
 Operator : TP/JJ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE102714

Quant Time: Oct 28 10:23:36 2014
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-PAH-BE102714.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 28 10:16:06 2014
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)

1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	152#	0.00
2	1,4-Dioxane	0.709	0.775	-9.3	162#	0.09
3 I	Naphthalene-d8	1.000	1.000	0.0	145	0.00
4 S	Nitrobenzene-d5	0.377	0.409	-8.5	160#	0.00
5	Naphthalene	1.046	1.131	-8.1	155#	0.00
6	2-Methylnaphthalene	0.678	0.715	-5.5	149	0.00
7 I	Acenaphthene-d10	1.000	1.000	0.0	125	0.00
8 S	2-Fluorobiphenyl	1.466	1.562	-6.5	142	0.00
9	Acenaphthylene	8.674	9.056	-4.4	142	0.00
10	Acenaphthene	1.303	1.295	0.6	137	-0.01
11	Fluorene	1.510	1.482	1.9	137	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	121	0.00
13	Phenanthrene	4.999	5.269	-5.4	140	0.00
14	Anthracene	4.268	4.138	3.0	137	0.00
15	Fluoranthene	0.981	0.917	6.5	137	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	131	0.00
17	Pyrene	1.729	1.578	8.7	139	0.00
18 S	Terphenyl-d14	1.060	0.976	7.9	139	0.00
19	Benzo(a)anthracene	5.434	4.767	12.3	150#	0.00
20	Chrysene	5.476	4.727	13.7	150	0.00
21	Indeno(1,2,3-cd)pyrene	5.406	4.799	11.2	177#	0.00
22 I	Perylene-d12	1.000	1.000	0.0	152#	0.00
23	Benzo(b)fluoranthene	1.525	1.245	18.4	168#	0.00
24	Benzo(k)fluoranthene	1.569	1.362	13.2	172#	0.00
25 C	Benzo(a)pyrene	1.303	1.168	10.4	174#	0.00
26	Dibenzo(a,h)anthracene	1.267	1.085	14.4	175#	0.00
27	Benzo(g,h,i)perylene	5.579	4.752	14.8	173#	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0