

Data Path : Z:\HPCHEM1\BNA E\DATA\BE102715\
 Data File : BE091052.D
 Acq On : 27 Oct 2015 13:31
 Operator : UM/NP
 Sample : SSTD3.294
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTD3.294

Quant Time: Oct 27 14:14:15 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM02.2-EPA-SIM-BE102715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 27 14:33:17 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.91	152	4094	0.40	ng/ul	0.00
4) Naphthalene-d8	9.58	136	19993	0.40	ng/ul	0.00
8) Acenaphthene-d10	13.41	164	10395	0.40	ng/ul	0.00
12) Phenanthrene-d10	16.14	188	27709	0.40	ng/ul	0.00
18) Chrysene-d12	20.29	240	35933	0.40	ng/ul	0.00
22) Perylene-d12	22.19	264	29383	0.40	ng/ul	0.00
System Monitoring Compounds						
2) 1,4-Dioxane-d8	0.00	96	0d	0.00	ng/uL	
6) 2-Methylnaphthalene-d10	0.00	152	0d	0.00	ng/ul	
16) Fluoranthene-d10	0.00	212	0d	0.00	ng/ul	
Target Compounds						
13) Pentachlorophenol	15.89	266	12177	5.17	ng/ul	Ovalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE102715\
Data File : BE091052.D
Acq On : 27 Oct 2015 13:31
Operator : UM/NP
Sample : SSTD3.294
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTD3.294

Quant Time: Oct 27 14:14:15 2015
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM02.2-EPA-SIM-BE102715.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 27 14:33:17 2015
Response via : Initial Calibration

