

Data Path : Z:\HPCHEM1\BNA E\Data\Be102715\
 Data File : BE091053.D
 Acq On : 27 Oct 2015 15:24
 Operator : UM/NP
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.483

Quant Time: Oct 28 02:57:01 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM02.2-EPA-SIM-BE102715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 28 03:51:36 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
2 SURR	1,4-Dioxane-d8	0.288	0.308	-6.9	100	0.00
3	1,4-Dioxane	0.325	0.332	-2.2	100	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	100	0.00
5	Naphthalene	0.986	1.015	-2.9	100	0.00
6 SURR	2-Methylnaphthalene-d10	0.546	0.549	-0.5	100	0.00
7	2-Methylnaphthalene	0.643	0.651	-1.2	100	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	100	0.00
9	Acenaphthylene	1.998	2.056	-2.9	100	0.00
10 C	Acenaphthene	1.368	1.399	-2.3	100	0.00
11	Fluorene	1.757	1.886	-7.3	100	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	100	0.00
13	Pentachlorophenol	0.038	0.030	21.1#	100	0.00
14	Phenanthrene	4.619	4.683	-1.4	100	0.00
15	Anthracene	4.200	4.073	3.0	100	0.00
16 SURR	Fluoranthene-d10	1.127	1.059	6.0	100	0.00
17 C	Fluoranthene	5.406	5.180	4.2	100	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	100	0.00
19	Pyrene	4.544	4.584	-0.9	100	0.00
20	Benzo(a)anthracene	1.027	1.000	2.6	100	0.00
21	Chrysene	1.183	1.197	-1.2	100	0.00
22 I	Perylene-d12	1.000	1.000	0.0	100	0.00
23	Benzo(b)fluoranthene	1.249	1.277	-2.2	100	0.00
24	Benzo(k)fluoranthene	1.281	1.342	-4.8	100	0.00
25 C	Benzo(a)pyrene	1.150	1.170	-1.7	100	0.00
26	Indeno(1,2,3-cd)pyrene	4.847	5.003	-3.2	100	0.00
27	Dibenzo(a,h)anthracene	1.155	1.163	-0.7	100	0.00
28	Benzo(q,h,i)perylene	5.083	5.145	-1.2	100	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0