

Data Path : Z:\HPCHEM1\BNA E\DATA\BE102715\
 Data File : BE091060.D
 Acq On : 27 Oct 2015 22:01
 Operator : UM/NP
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.484

Quant Time: Oct 28 03:01:04 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM02.2-EPA-SIM-BE102715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 28 03:57:22 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	122	-0.02
2 SURR	1,4-Dioxane-d8	0.288	0.257	10.8	102	-0.05
3	1,4-Dioxane	0.325	0.281	13.5	104	-0.05
4 I	Naphthalene-d8	1.000	1.000	0.0	138	0.00
5	Naphthalene	0.986	1.023	-3.8	139	0.00
6 SURR	2-Methylnaphthalene-d10	0.546	0.592	-8.4	148	-0.01
7	2-Methylnaphthalene	0.643	0.697	-8.4	147	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	145	0.00
9	Acenaphthylene	1.998	2.032	-1.7	143	0.00
10 C	Acenaphthene	1.368	1.356	0.9	141	0.00
11	Fluorene	1.757	1.585	9.8	122	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	108	0.00
13	Pentachlorophenol	0.038	0.081	-113.2#	289#	0.00
14	Phenanthrene	4.619	4.635	-0.3	107	0.00
15	Anthracene	4.200	4.578	-9.0	122	0.01
16 SURR	Fluoranthene-d10	1.127	1.106	1.9	113	0.00
17 C	Fluoranthene	5.406	5.219	3.5	109	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	107	0.00
19	Pyrene	4.544	4.966	-9.3	116	0.00
20	Benzo(a)anthracene	1.027	1.179	-14.8	126	0.01
21	Chrysene	1.183	1.165	1.5	104	0.00
22 I	Perylene-d12	1.000	1.000	0.0	121	0.01
23	Benzo(b)fluoranthene	1.249	1.255	-0.5	119	0.00
24	Benzo(k)fluoranthene	1.281	1.183	7.7	107	0.00
25 C	Benzo(a)pyrene	1.150	1.175	-2.2	121	0.01
26	Indeno(1,2,3-cd)pyrene	4.847	4.719	2.6	114	0.02
27	Dibenzo(a,h)anthracene	1.155	1.220	-5.6	127	0.01
28	Benzo(q,h,i)perylene	5.083	4.856	4.5	114	0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0