

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE102819\
 Data File : BE100681.D
 Acq On : 28 Oct 2019 19:43
 Operator : JU
 Sample : K5546-01
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 GW809-8.0-13.0-102419

Manual Integrations
 APPROVED

mohammad
 10/30/2019 4:52:07 PM

Quant Time: Oct 29 14:01:08 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE101019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 23 16:49:33 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.85	152	3313	0.40	ng	0.00
7) Naphthalene-d8	10.63	136	14078	0.40	ng	0.00
13) Acenaphthene-d10	14.47	164	8972	0.40	ng	0.00
19) Phenanthrene-d10	17.19	188	21143	0.40	ng	-0.01
27) Chrysene-d12	21.35	240	30505	0.40	ng	0.00
34) Perylene-d12	23.82	264	35068	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.44	112	2821	0.28	ng	0.00
5) Phenol-d6	7.01	99	2826	0.20	ng	0.00
8) Nitrobenzene-d5	8.99	82	4346	0.50	ng	0.00
11) 2-Methylnaphthalene-d10	12.22	152	8359	0.38	ng	0.00
14) 2,4,6-Tribromophenol	15.94	330	1355	0.57	ng	0.00
15) 2-Fluorobiphenyl	13.10	172	13587	0.42	ng	0.00
25) Fluoranthene-d10	19.21	212	121764	0.45	ng	0.00
29) Terphenyl-d14	19.82	244	32623	0.49	ng	0.00
Target Compounds						
22) Pentachlorophenol	16.84	266	147m	0.045	ng	Qvalue
23) Phenanthrene	17.23	178	1746	0.030	ng	# 90
32) Bis(2-ethylhexyl)phthalate	21.28	149	61838	0.297	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE102819\
Data File : BE100681.D
Acq On : 28 Oct 2019 19:43
Operator : JU
Sample : K5546-01
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
GW809-8.0-13.0-102419

Manual Integrations
APPROVED

mohammad
10/30/2019 4:52:07 PM

Quant Time: Oct 29 14:01:08 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE101019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 23 16:49:33 2019
Response via : Initial Calibration

