

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE102819\
 Data File : BE100675.D
 Acq On : 28 Oct 2019 16:02
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 29 11:52:15 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE101019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 23 16:49:33 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	47	0.00
2	1,4-Dioxane	0.400	0.480	-20.0	57	0.00
3	n-Nitrosodimethylamine	0.400	0.419	-4.7	54	0.00
4 S	2-Fluorophenol	0.400	0.413	-3.2	52	0.00
5 S	Phenol-d6	0.400	0.404	-1.0	51	0.00
6	bis(2-Chloroethyl)ether	0.400	0.457	-14.2	58	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	50	0.00
8 S	Nitrobenzene-d5	0.400	0.519	-29.8#	74	0.00
9	Naphthalene	0.400	0.446	-11.5	57	0.00
10	Hexachlorobutadiene	0.400	0.459	-14.8	60	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.433	-8.2	56	0.00
12	2-Methylnaphthalene	0.400	0.445	-11.2	58	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	50	0.00
14 S	2,4,6-Tribromophenol	0.400	0.437	-9.2	62	0.00
15 S	2-Fluorobiphenyl	0.400	0.432	-8.0	55	0.00
16	Acenaphthylene	0.400	0.424	-6.0	55	0.00
17	Acenaphthene	0.400	0.438	-9.5	56	0.00
18	Fluorene	0.400	0.435	-8.7	56	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	51	-0.01
20	4-Bromophenyl-phenylether	0.400	0.440	-10.0	59	0.00
21	Hexachlorobenzene	0.400	0.446	-11.5	60	0.00
22	Pentachlorophenol	0.400	0.381	4.8	55	0.00
23	Phenanthrene	0.400	0.445	-11.2	57	-0.01
24	Anthracene	0.400	0.414	-3.5	55	0.00
25 SURR	Fluoranthene-d10	0.400	0.446	-11.5	60	0.00
26	Fluoranthene	0.400	0.454	-13.5	59	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	51	0.00
28	Pvrene	0.400	0.451	-12.7	57	0.00
29 S	Terphenyl-d14	0.400	0.443	-10.7	59	0.00
30	Benzo(a)anthracene	0.400	0.440	-10.0	56	0.00
31	Chrysene	0.400	0.457	-14.2	58	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.360	10.0	46	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.440	-10.0	58	-0.01
34 I	Perylene-d12	0.400	0.400	0.0	49	0.00
35	Benzo(b)fluoranthene	0.400	0.429	-7.2	57	0.00
36	Benzo(k)fluoranthene	0.400	0.434	-8.5	57	0.00
37 C	Benzo(a)pyrene	0.400	0.420	-5.0	56	0.00
38	Dibenzo(a,h)anthracene	0.400	0.431	-7.7	59	0.00
39	Benzo(a,h,i)perylene	0.400	0.443	-10.7	58	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE102819\
Data File : BE100675.D
Acq On : 28 Oct 2019 16:02
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 29 11:52:15 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE101019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 23 16:49:33 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				