

Data Path : Z:\HPCHEM1\BNA E\DATA\BE103017\
 Data File : BE094649.D
 Acq On : 30 Oct 2017 18:53
 Operator : SJ/JU
 Sample : I6036-02
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :

Quant Time: Oct 31 01:16:03 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE102317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 23 22:40:47 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.90	152	1406	0.40	ng	0.00
7) Naphthalene-d8	10.69	136	8273	0.40	ng	-0.02
13) Acenaphthene-d10	14.51	164	7925	0.40	ng	0.00
19) Phenanthrene-d10	17.23	188	8560	0.40	ng	0.00
27) Chrysene-d12	21.43	240	9329	0.40	ng	0.00
34) Perylene-d12	23.96	264	8766	0.40	ng	0.01

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Fluorophenol	5.54	112	606	0.13	ng	0.01
5) Phenol-d6	7.12	99	715	0.10	ng	-0.01
8) Nitrobenzene-d5	9.06	82	1773	0.27	ng	-0.02
11) 2-Methylnaphthalene-d10	12.27	152	6633	0.50	ng	-0.01
14) 2,4,6-Tribromophenol	16.03	330	434	0.20	ng	0.00
15) 2-Fluorobiphenyl	13.13	172	7389	0.26	ng	-0.02
25) Fluoranthene-d10	19.28	212	45104	0.43	ng	0.00
29) Terphenyl-d14	19.86	244	8677	0.49	ng	0.00


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) 1,4-Dioxane	3.41	88	113	0.05	ng	# 1
3) n-Nitrosodimethylamine	3.78	42	99	0.04	ng	# 9
9) Naphthalene	10.72	128	36710	0.39	ng	# 58
12) 2-Methylnaphthalene	12.35	142	1379	0.09	ng	# 6
16) Acenaphthylene	14.23	152	16231	0.39	ng	# 1
17) Acenaphthene	14.57	154	36468	1.36	ng	# 61
18) Fluorene	15.56	166	95759	2.78	ng	# 59
20) 4-Bromophenyl-phenylether	16.39	248	561	0.12	ng	# 1
22) Pentachlorophenol	17.04	266	203	0.39	ng	# 74
23) Phenanthrene	17.30	178	99864	5.08	ng	# 61
24) Anthracene	17.43	178	65398	2.60	ng	# 1
26) Fluoranthene	19.31	202	9951	0.29	ng	# 87
28) Pyrene	19.67	202	48862	1.50	ng	# 70
30) Benzo(a)anthracene	21.44	228	5379	0.18	ng	# 70
31) Chrysene	21.44	228	5263	0.17	ng	# 70
32) Bis(2-ethylhexyl)phthalate	21.28	149	21029	0.28	ng	97
33) Indeno(1,2,3-cd)pyrene	26.67	276	618	0.02	ng	# 87
35) Benzo(b)fluoranthene	23.19	252	1728	0.06	ng	94
36) Benzo(k)fluoranthene	23.19	252	1681	0.06	ng	94
37) Benzo(a)pyrene	23.74	252	1218	0.04	ng	94
39) Benzo(g,h,i)perylene	27.52	276	626	0.03	ng	# 82


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE103017\
Data File : BE094649.D
Acq On : 30 Oct 2017 18:53
Operator : SJ/JU
Sample : I6036-02
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :

Quant Time: Oct 31 01:16:03 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE102317.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 23 22:40:47 2017
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.90 min Scan# 401
Delta R.T. -0.00 min
Lab File: BE094649.D
Acq: 30 Oct 2017 18:53

Instrument :
BNA_E
ClientSampleId :

Tot Ion:152 Resp: 1406
Ion Ratio Lower Upper
152 100
150 149.3 117.2 175.8
115 76.0 57.0 85.6




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#2


1,4-Dioxane
Concen: 0.05 ng
RT: 3.41 min Scan# 44
Delta R.T. -0.01 min
Lab File: BE094649.D
Acq: 30 Oct 2017 18:53

Tgt Ion: 88 Resp: 113
Ion Ratio Lower Upper
88 100
58 0.0 63.2 94.8#
43 203.5 41.2 61.8#



Abundance Ion 88.00 (87.70 to 88.70): BE0
Ion 58.00 (57.70 to 58.70): BE0
Ion 43.00 (42.70 to 43.70): BE0






#3




n-Nitrosodimethylamine
 Concen: 0.04 ng
 RT: 3.78 min Scan# 73
 Delta R.T. 0.02 min
 Lab File: BE094649.D
 Acq: 30 Oct 2017 18:53

Instrument :
 BNA_E
 ClientSampleId :

Tot Ion: 42 Resp: 99
 Ion Ratio Lower Upper
 42 100
 74 12.1 100.3 150.5#
 44 0.0 17.0 25.4#




Abundance Ion 42.00 (41.70 to 42.70): BE0
 Ion 74.00 (73.70 to 74.70): BE0
 Ion 44.00 (43.70 to 44.70): BE0





#4


2-Fluorophenol
 Concen: 0.13 ng
 RT: 5.54 min Scan# 213
 Delta R.T. 0.01 min
 Lab File: BE094649.D
 Acq: 30 Oct 2017 18:53

Tgt Ion: 112 Resp: 606
 Ion Ratio Lower Upper
 112 100
 64 77.7 60.1 90.1
 63 155.6 116.8 175.2



Abundance Ion 112.00 (111.70 to 112.70): E
 Ion 64.00 (63.70 to 64.70): BE0
 Ion 63.00 (62.70 to 63.70): BE0





#5

Phenol-d6

Concen: 0.10 ng

RT: 7.12 min Scan# 339

Delta R.T. -0.01 min

Lab File: BE094649.D

Acq: 30 Oct 2017 18:53

Instrument :

BNA_E

ClientSampleId :


Tot Ion: 99 Resp: 715

Ion Ratio Lower Upper




99 100

42 0.0 20.2 30.2#

71 111.9 28.4 42.6#



Abundance Ion 99.00 (98.70 to 99.70): BE0
Ion 42.00 (41.70 to 42.70): BE0
Ion 71.00 (70.70 to 71.70): BE0



#7

Naphthalene-d8

Concen: 0.40 ng

RT: 10.69 min Scan# 576

Delta R.T. -0.02 min

Lab File: BE094649.D

Acq: 30 Oct 2017 18:53

Tgt Ion: 136 Resp: 8273


Ion Ratio Lower Upper

136 100



137 27.0 9.5 14.3#


54 135.6 29.1 43.7#

68 50.4 6.2 9.2#



Abundance Ion 136.00 (135.70 to 136.70): E
Ion 137.00 (136.70 to 137.70): E
Ion 54.00 (53.70 to 54.70): BE0
Ion 68.00 (67.70 to 68.70): BE0





#8

Nitrobenzene-d5

Concen: 0.27 ng

RT: 9.06 min Scan# 479

Delta R.T. -0.02 min

Lab File: BE094649.D

Acq: 30 Oct 2017 18:53

Instrument :

BNA_E

ClientSampleId :


Tot Ion: 82 Resp: 1773

Ion Ratio Lower Upper

82 100

128 310.9 150.0 225.0#




54 291.8 154.2 231.4#



Abundance Ion 82.00 (81.70 to 82.70): BE0

Ion 128.00 (127.70 to 128.70): BE0

Ion 54.00 (53.70 to 54.70): BE0



#9

Naphthalene

Concen: 0.39 ng

RT: 10.72 min Scan# 578

Delta R.T. -0.02 min

Lab File: BE094649.D

Acq: 30 Oct 2017 18:53


Tgt Ion: 128 Resp: 36710

Ion Ratio Lower Upper

128 100

129 23.7 2.6 3.8#



127 11.5 2.8 4.2#




Abundance Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E





#11

2-Methylnaphthalene-d10

Concen: 0.50 ng

RT: 12.27 min Scan# 820

Delta R.T. -0.01 min

Lab File: BE094649.D

Acq: 30 Oct 2017 18:53

Instrument :

BNA_E





ClientSampleId :

Tot Ion:152 Resp: 6633

Ion Ratio Lower Upper

152 100

151 0.0 15.4 23.0#



#12

2-Methylnaphthalene

Concen: 0.09 ng

RT: 12.35 min Scan# 840

Delta R.T. -0.01 min

Lab File: BE094649.D

Acq: 30 Oct 2017 18:53




Tgt Ion:142 Resp: 1379


Ion Ratio Lower Upper

142 100

141 114.5 70.4 105.6#

115 186.6 31.7 47.5#





#13

Acenaphthene-d10

Concen: 0.40 ng

RT: 14.51 min Scan# 1110

Delta R.T. -0.00 min


Lab File: BE094649.D

Acq: 30 Oct 2017 18:53

Instrument :

BNA_E

ClientSampled :






Tot Ion:164 Resp: 7925

Ion Ratio Lower Upper

164 100

162 83.8 83.1 124.7

160 181.0 37.1 55.7#



#14

2,4,6-Tribromophenol

Concen: 0.20 ng

RT: 16.03 min Scan# 1205

Delta R.T. -0.00 min

Lab File: BE094649.D

Acq: 30 Oct 2017 18:53




Tgt Ion:330 Resp: 434


Ion Ratio Lower Upper

330 100

332 112.2 152.7 229.1#

141 770.3 33.7 50.5#









#15
2-Fluorobiphenyl
Concen: 0.26 ng
RT: 13.13 min Scan# 1017
Delta R.T. -0.02 min
Lab File: BE094649.D
Acq: 30 Oct 2017 18:53




Instrument :
BNA_E
ClientSampled :


Tot Ion:172 Resp: 7389
Ion Ratio Lower Upper
172 100
171 37.6 29.9 44.9
170 24.5 21.8 32.8



#16
Acenaphthylene
Concen: 0.39 ng
RT: 14.23 min Scan# 1091
Delta R.T. -0.02 min
Lab File: BE094649.D
Acq: 30 Oct 2017 18:53

Tgt Ion:152 Resp: 16231
Ion Ratio Lower Upper
152 100
151 42.6 15.7 23.5#
153 92.0 10.5 15.7#





#17

Acenaphthene

Concen: 1.36 ng

RT: 14.57 min Scan# 1114

Delta R.T. -0.02 min

Lab File: BE094649.D

Acq: 30 Oct 2017 18:53

Instrument :

BNA_E

ClientSampleId :


Tot Ion:154 Resp: 36468

Ion Ratio Lower Upper

154 100

153 142.3 89.2 133.8#




152 94.2 42.0 63.0#



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#18

Fluorene

Concen: 2.78 ng

RT: 15.56 min Scan# 1181

Delta R.T. -0.00 min

Lab File: BE094649.D

Acq: 30 Oct 2017 18:53


Tgt Ion:166 Resp: 95759

Ion Ratio Lower Upper

166 100

165 119.8 77.8 116.6#



167 0.0 42.4 63.6#




Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.23 min Scan# 1242
Delta R.T. -0.00 min
Lab File: BE094649.D
Acq: 30 Oct 2017 18:53




Instrument :
BNA_E
ClientSampled :


Tot Ion:188 Resp: 8560
Ion Ratio Lower Upper
188 100
94 85.9 3.9 5.9#
80 79.9 3.6 5.4#



#20
4-Bromophenyl-phenylether
Concen: 0.12 ng
RT: 16.39 min Scan# 1216
Delta R.T. -0.03 min
Lab File: BE094649.D
Acq: 30 Oct 2017 18:53

Tgt Ion:248 Resp: 561
Ion Ratio Lower Upper
248 100
250 91.9 62.7 94.1
141 2210.7 48.2 72.2#





#22

Pentachlorophenol

Concen: 0.39 ng

RT: 17.04 min Scan# 1236

Delta R.T. 0.03 min

Lab File: BE094649.D

Acq: 30 Oct 2017 18:53

Instrument :

BNA_E

ClientSampleId :


Tot Ion:266 Resp: 203

Ion Ratio Lower Upper

266 100

264 107.4 72.5 108.7




268 60.5 73.8 110.6#



Abundance Ion 266.00 (265.70 to 266.70): E

Ion 264.00 (263.70 to 264.70): E

Ion 268.00 (267.70 to 268.70): E



#23

Phenanthrene

Concen: 5.08 ng

RT: 17.30 min Scan# 1244

Delta R.T. -0.00 min

Lab File: BE094649.D

Acq: 30 Oct 2017 18:53


Tgt Ion:178 Resp: 99864

Ion Ratio Lower Upper

178 100

176 0.0 15.1 22.7#



179 0.0 11.8 17.6#




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E





#24

Anthracene

Concen: 2.60 ng

RT: 17.43 min Scan# 1248

Delta R.T. 0.03 min

Lab File: BE094649.D

Acq: 30 Oct 2017 18:53

Instrument :

BNA_E

ClientSampled :


Tot Ion:178 Resp: 65398

Ion Ratio Lower Upper

178 100

176 31.7 12.8 19.2#




179 415.6 13.0 19.6#



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E



#25

Fluoranthene-d10

Concen: 0.43 ng

RT: 19.28 min Scan# 1432

Delta R.T. 0.00 min

Lab File: BE094649.D

Acq: 30 Oct 2017 18:53


Tgt Ion:212 Resp: 45104

Ion Ratio Lower Upper

212 100

106 4.6 2.8 4.2#



104 1.4 1.6 2.4#




Abundance Ion 212.00 (211.70 to 212.70): E

Ion 106.00 (105.70 to 106.70): E

Ion 104.00 (103.70 to 104.70): E









#26
Fluoranthene
Concen: 0.29 ng
RT: 19.31 min Scan# 1436
Delta R.T. -0.00 min
Lab File: BE094649.D
Acq: 30 Oct 2017 18:53




Instrument :
BNA_E
ClientSampled :


Tot Ion:202 Resp: 9951
Ion Ratio Lower Upper
202 100
101 19.9 9.8 14.8#
203 20.4 13.7 20.5



#27
Chrysene-d12
Concen: 0.40 ng
RT: 21.43 min Scan# 1655
Delta R.T. 0.00 min
Lab File: BE094649.D
Acq: 30 Oct 2017 18:53

Tgt Ion:240 Resp: 9329
Ion Ratio Lower Upper
240 100
120 20.0 5.5 8.3#
236 30.9 20.7 31.1





#28

Pvrene

Concen: 1.50 ng

RT: 19.67 min Scan# 1485

Delta R.T. -0.00 min

Lab File: BE094649.D

Acq: 30 Oct 2017 18:53

Instrument :

BNA_E

ClientSampleId :





Tot Ion:202 Resp: 48862

Ion Ratio Lower Upper

202 100

200 18.9 16.6 25.0

203 43.9 13.8 20.8#



#29

Terphenyl-d14

Concen: 0.49 ng

RT: 19.86 min Scan# 1510

Delta R.T. -0.01 min

Lab File: BE094649.D

Acq: 30 Oct 2017 18:53




Tgt Ion:244 Resp: 8677


Ion Ratio Lower Upper

244 100

212 90.2 25.4 38.0#

122 32.1 8.1 12.1#





#30

Benzo(a)anthracene

Concen: 0.18 ng

RT: 21.44 min Scan# 1656

Delta R.T. 0.03 min

Lab File: BE094649.D

Acq: 30 Oct 2017 18:53

Instrument :

BNA_E

ClientSampleId :





Tot Ion:228 Resp: 5379

Ion Ratio Lower Upper

228 100

226 53.9 40.0 60.0

229 87.4 40.0 60.0#



#31

Chrysene

Concen: 0.17 ng

RT: 21.44 min Scan# 1656

Delta R.T. -0.02 min

Lab File: BE094649.D

Acq: 30 Oct 2017 18:53




Tgt Ion:228 Resp: 5263


Ion Ratio Lower Upper

228 100

226 53.9 40.0 60.0

229 87.4 40.0 60.0#









#32
 Bis(2-ethylhexyl)phthalate
 Concen: 0.28 ng
 RT: 21.28 min Scan# 1642
 Delta R.T. -0.01 min
 Lab File: BE094649.D
 Acq: 30 Oct 2017 18:53




Instrument :
 BNA_E
 ClientSampleId :




Tot Ion:149 Resp: 21029
 Ion Ratio Lower Upper
 149 100
 167 8.0 5.4 8.0
 279 1.1 0.7 1.1



#33
 Indeno(1,2,3-cd)pyrene
 Concen: 0.02 ng
 RT: 26.67 min Scan# 2723
 Delta R.T. 0.04 min
 Lab File: BE094649.D
 Acq: 30 Oct 2017 18:53

Tgt Ion:276 Resp: 618
 Ion Ratio Lower Upper
 276 100
 138 20.4 22.5 33.7#
 277 30.4 19.9 29.9#









#34
 Pervlene-d12
 Concen: 0.40 ng
 RT: 23.96 min Scan# 2132
 Delta R.T. 0.01 min
 Lab File: BE094649.D
 Acq: 30 Oct 2017 18:53


Instrument :
 BNA_E
 ClientSampleId :


Tot Ion:264 Resp: 8766
 Ion Ratio Lower Upper
 264 100
 260 27.7 20.5 30.7
 265 34.7 22.8 34.2#



#35
 Benzo(b)fluoranthene
 Concen: 0.06 ng
 RT: 23.19 min Scan# 1962
 Delta R.T. 0.01 min
 Lab File: BE094649.D
 Acq: 30 Oct 2017 18:53

Tgt Ion:252 Resp: 1728
 Ion Ratio Lower Upper
 252 100
 253 61.8 48.0 72.0
 125 77.2 56.0 84.0





#36

Benzo(k)fluoranthene

Concen: 0.06 ng

RT: 23.19 min Scan# 1962

Delta R.T. -0.03 min

Lab File: BE094649.D

Acq: 30 Oct 2017 18:53

Instrument :

BNA_E

ClientSampleId :


Tot Ion:252 Resp: 1681

Ion Ratio Lower Upper

252 100

253 61.8 48.0 72.0




125 77.2 56.0 84.0



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#37

Benzo(a)pyrene

Concen: 0.04 ng

RT: 23.74 min Scan# 2083

Delta R.T. -0.10 min

Lab File: BE094649.D

Acq: 30 Oct 2017 18:53


Tgt Ion:252 Resp: 1218

Ion Ratio Lower Upper

252 100

253 69.5 52.0 78.0



125 90.5 68.0 102.0




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E





#39

Benzo(a,h,i)perylene

Concen: 0.03 ng

RT: 27.52 min Scan# 2909

Delta R.T. 0.05 min


Lab File: BE094649.D

Acq: 30 Oct 2017 18:53

Instrument :

BNA_E

ClientSampleId :



Tot Ion: 276 Resp: 626

Ion Ratio Lower Upper

276 100

277 68.8 52.0 78.0

138 79.6 44.0 66.0

