

Data Path : Z:\HPCHEM1\BNA E\DATA\BE103017\
 Data File : BE094651.D
 Acq On : 30 Oct 2017 20:05
 Operator : SJ/JU
 Sample : I6036-04
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 MW-5

Quant Time: Oct 31 01:44:05 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE102317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 23 22:40:47 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.90	152	1243	0.40	ng	0.00
7) Naphthalene-d8	10.69	136	7091	0.40	ng	-0.02
13) Acenaphthene-d10	14.51	164	4181	0.40	ng	0.00
19) Phenanthrene-d10	17.24	188	8315	0.40	ng	0.00
27) Chrysene-d12	21.43	240	9098	0.40	ng	0.00
34) Perylene-d12	23.97	264	8489	0.40	ng	0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.54	112	716	0.17	ng	0.01
5) Phenol-d6	7.14	99	678	0.11	ng	0.00
8) Nitrobenzene-d5	9.09	82	2272	0.40	ng	0.02
11) 2-Methylnaphthalene-d10	12.28	152	4832	0.42	ng	-0.01
14) 2,4,6-Tribromophenol	16.03	330	526	0.46	ng	0.00
15) 2-Fluorobiphenyl	13.13	172	6687	0.45	ng	-0.01
25) Fluoranthene-d10	19.28	212	48216	0.49	ng	0.00
29) Terphenyl-d14	19.86	244	8327	0.49	ng	0.00
Target Compounds						
17) Acenaphthene	14.57	154	1309	0.092	ng	# 67
23) Phenanthrene	17.30	178	3647	0.124	ng	# 61
32) Bis(2-ethylhexyl)phthalate	21.28	149	3841	0.053	ng	# 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE103017\
Data File : BE094651.D
Acq On : 30 Oct 2017 20:05
Operator : SJ/JU
Sample : I6036-04
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
MW-5

Quant Time: Oct 31 01:44:05 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE102317.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 23 22:40:47 2017
Response via : Initial Calibration

