

Data Path : Z:\HPCHEM1\BNA_E\Data\BE103017\
 Data File : BE094644.D
 Acq On : 30 Oct 2017 15:54
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 31 19:37:51 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE102317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 23 22:40:47 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	105	0.00
2	1,4-Dioxane	0.677	0.602	11.1	80	0.00
3	n-Nitrosodimethylamine	0.648	0.528	18.5	85	0.00
4 S	2-Fluorophenol	1.370	1.196	12.7	91	0.01
5 S	Phenol-d6	2.069	2.027	2.0	105	0.01
6	bis(2-Chloroethyl)ether	1.600	1.435	10.3	95	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	113	0.00
8 S	Nitrobenzene-d5	0.319	0.279	12.5	103	0.00
9	Naphthalene	4.527	3.923	13.3	96	0.00
10	Hexachlorobutadiene	0.164	0.157	4.3	107	0.00
11 SURR	2-Methylnaphthalene-d10	0.642	0.558	13.1	96	0.00
12	2-Methylnaphthalene	0.768	0.662	13.8	97	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	102	0.02
14 S	2,4,6-Tribromophenol	0.110	0.100	9.1	96	0.03
15 S	2-Fluorobiphenyl	1.410	1.309	7.2	93	0.00
16	Acenaphthylene	2.112	1.871	11.4	91	0.00
17	Acenaphthene	1.356	1.218	10.2	90	0.00
18	Fluorene	1.738	1.506	13.3	88	0.02
19 I	Phenanthrene-d10	1.000	1.000	0.0	96	0.03
20	4-Bromophenyl-phenylether	0.218	0.184	15.6	83	0.00
21	Hexachlorobenzene	0.191	0.293	-53.4#	128	0.00
22	Pentachlorophenol	0.024	0.026	-8.3	102	0.00
23	Phenanthrene	1.141	1.216	-6.6	102	0.00
24	Anthracene	1.176	1.062	9.7	85	0.00
25 SURR	Fluoranthene-d10	4.676	5.104	-9.2	91	0.00
26	Fluoranthene	1.420	1.545	-8.8	91	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	100	0.00
28	Pyrene	1.399	1.255	10.3	91	0.00
29 S	Terphenyl-d14	0.754	0.677	10.2	91	0.00
30	Benzo(a)anthracene	1.302	1.181	9.3	91	0.01
31	Chrysene	1.322	1.205	8.9	94	0.01
32	Bis(2-ethylhexyl)phthalate	3.193	2.947	7.7	101	0.00
33	Indeno(1,2,3-cd)pyrene	1.222	0.985	19.4	82	0.05
34 I	Perylene-d12	1.000	1.000	0.0	98	0.02
35	Benzo(b)fluoranthene	1.294	1.197	7.5	94	0.01
36	Benzo(k)fluoranthene	1.379	1.261	8.6	91	0.02
37 C	Benzo(a)pyrene	1.255	1.163	7.3	91	0.02
38	Dibenzo(a,h)anthracene	1.127	0.970	13.9	87	0.04
39	Benzo(g,h,i)perylene	1.055	0.856	18.9	81	0.06

Data Path : Z:\HPCHEM1\BNA_E\Data\BE103017\
Data File : BE094644.D
Acq On : 30 Oct 2017 15:54
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 31 19:37:51 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE102317.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 23 22:40:47 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			