

Data Path : Z:\HPCHEM1\BNA_E\Data\BE103114\
 Data File : BE088433.D
 Acq On : 31 Oct 2014 21:29
 Operator : TP/JJ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Oct 31 23:47:57 2014
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-PAH-BE103114.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Nov 01 00:42:07 2014
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
2	1,4-Dioxane	0.654	0.682	-4.3	100	0.00
3 I	Naphthalene-d8	1.000	1.000	0.0	99	0.00
4 S	Nitrobenzene-d5	0.367	0.381	-3.8	99	0.00
5	Naphthalene	1.035	1.064	-2.8	98	0.00
6	2-Methylnaphthalene	0.641	0.667	-4.1	98	0.00
7 I	Acenaphthene-d10	1.000	1.000	0.0	98	0.00
8 S	2-Fluorobiphenyl	1.409	1.460	-3.6	97	0.00
9	Acenaphthylene	8.098	8.312	-2.6	98	0.00
10	Acenaphthene	1.201	1.212	-0.9	97	0.00
11	Fluorene	1.273	1.311	-3.0	95	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	94	0.00
13	Phenanthrene	4.543	5.147	-13.3	93	0.00
14	Anthracene	3.669	3.819	-4.1	95	0.00
15	Fluoranthene	0.737	0.761	-3.3	92	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	93	0.00
17	Pyrene	1.631	1.686	-3.4	91	0.00
18 S	Terphenyl-d14	0.919	0.945	-2.8	92	0.00
19	Benzo(a)anthracene	4.707	4.819	-2.4	93	0.00
20	Chrysene	4.620	4.804	-4.0	94	0.00
21	Indeno(1,2,3-cd)pyrene	4.306	4.437	-3.0	95	0.00
22 I	Perylene-d12	1.000	1.000	0.0	96	0.00
23	Benzo(b)fluoranthene	1.096	1.136	-3.6	97	0.00
24	Benzo(k)fluoranthene	1.207	1.244	-3.1	94	0.00
25 C	Benzo(a)pyrene	1.053	1.086	-3.1	96	0.00
26	Dibenzo(a,h)anthracene	0.978	1.006	-2.9	96	0.00
27	Benzo(g,h,i)perylene	4.483	4.606	-2.7	97	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0