

Data Path : Z:\HPCHEM1\BNA_E\Data\BE103116\
 Data File : BE092481.D
 Acq On : 31 Oct 2016 14:52
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 01 11:03:46 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE102516.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 25 15:25:33 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	96	0.00
2	1,4-Dioxane	0.400	0.471	-17.7	118	0.00
3	n-Nitrosodimethylamine	0.400	0.480	-20.0	101	-0.01
4 S	2-Fluorophenol	0.400	0.412	-3.0	108	0.00
5 S	Phenol-d6	0.400	0.395	1.3	100	0.00
6	bis(2-Chloroethyl)ether	0.400	0.429	-7.2	96	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	98	0.00
8 S	Nitrobenzene-d5	0.400	0.466	-16.5	102	0.00
9	Naphthalene	0.400	0.389	2.8	95	0.00
10	Hexachlorobutadiene	0.400	0.397	0.8	96	0.00
11	2-Methylnaphthalene	0.400	0.379	5.3	97	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	101	0.00
13 S	2,4,6-Tribromophenol	0.400	0.369	7.8	92	0.00
14 S	2-Fluorobiphenyl	0.400	0.405	-1.3	104	0.00
15	Acenaphthylene	0.400	0.383	4.3	100	0.00
16	Acenaphthene	0.400	0.396	1.0	101	0.00
17	Fluorene	0.400	0.386	3.5	101	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	96	0.00
19	Hexachlorobenzene	0.400	0.420	-5.0	105	0.00
20	Pentachlorophenol	0.400	0.441	-10.2	88	0.00
21	Phenanthrene	0.400	0.390	2.5	102	0.00
22	Anthracene	0.400	0.369	7.8	101	0.00
23	Fluoranthene	0.400	0.384	4.0	98	0.00
24 I	Chrysene-d12	5.000	5.000	0.0	97	0.00
25	Pyrene	0.400	0.369	7.8	96	-0.02
26 S	Terphenyl-d14	0.400	0.386	3.5	101	0.00
27	Benzo(a)anthracene	0.400	0.356	11.0	92	0.00
28	Chrysene	0.400	0.403	-0.8	101	0.00
29	Bis(2-ethylhexyl)phthalate	0.400	0.404	-1.0	86	0.00
30	Indeno(1,2,3-cd)pyrene	0.400	0.349	12.8	92	0.00
31 I	Perylene-d12	5.000	5.000	0.0	96	0.00
32	Benzo(b)fluoranthene	0.400	0.339	15.3	93	0.00
33	Benzo(k)fluoranthene	0.400	0.397	0.8	92	0.00
34 C	Benzo(a)pyrene	0.400	0.341	14.8	92	0.00
35	Dibenzo(a,h)anthracene	0.400	0.343	14.2	90	0.00
36	Benzo(g,h,i)perylene	0.400	0.364	9.0	92	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0