

Data Path : Z:\HPCHEM1\BNA_E\Data\BE103116\
 Data File : BE092481.D
 Acq On : 31 Oct 2016 14:52
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 01 11:03:46 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE102516.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 25 15:25:33 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
2	1,4-Dioxane	0.654	0.757	-15.7	118	0.00
3	n-Nitrosodimethylamine	0.600	0.606	-1.0	101	-0.01
4 S	2-Fluorophenol	0.758	0.771	-1.7	108	0.00
5 S	Phenol-d6	1.060	1.035	2.4	100	0.00
6	bis(2-Chloroethyl)ether	1.251	1.173	6.2	96	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	98	0.00
8 S	Nitrobenzene-d5	0.281	0.265	5.7	102	0.00
9	Naphthalene	1.081	1.052	2.7	95	0.00
10	Hexachlorobutadiene	0.166	0.165	0.6	96	0.00
11	2-Methylnaphthalene	0.698	0.662	5.2	97	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	101	0.00
13 S	2,4,6-Tribromophenol	0.098	0.083	15.3	92	0.00
14 S	2-Fluorobiphenyl	1.218	1.235	-1.4	104	0.00
15	Acenaphthylene	7.149	6.837	4.4	100	0.00
16	Acenaphthene	1.141	1.130	1.0	101	0.00
17	Fluorene	1.366	1.317	3.6	101	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	96	0.00
19	Hexachlorobenzene	0.224	0.235	-4.9	105	0.00
20	Pentachlorophenol	0.039	0.030	23.1	88	0.00
21	Phenanthrene	1.292	1.259	2.6	102	0.00
22	Anthracene	1.163	1.072	7.8	101	0.00
23	Fluoranthene	5.122	4.918	4.0	98	0.00
24 I	Chrysene-d12	1.000	1.000	0.0	97	0.00
25	Pyrene	4.358	4.015	7.9	96	-0.02
26 S	Terphenyl-d14	0.619	0.597	3.6	101	0.00
27	Benzo(a)anthracene	4.626	4.119	11.0	92	0.00
28	Chrysene	4.862	4.904	-0.9	101	0.00
29	Bis(2-ethylhexyl)phthalate	0.385	0.257	33.2#	86	0.00
30	Indeno(1,2,3-cd)pyrene	4.286	3.738	12.8	92	0.00
31 I	Perylene-d12	1.000	1.000	0.0	96	0.00
32	Benzo(b)fluoranthene	1.245	1.055	15.3	93	0.00
33	Benzo(k)fluoranthene	1.234	1.224	0.8	92	0.00
34 C	Benzo(a)pyrene	1.127	0.960	14.8	92	0.00
35	Dibenzo(a,h)anthracene	1.053	0.902	14.3	90	0.00
36	Benzo(g,h,i)perylene	4.540	4.136	8.9	92	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0