

Data Path : Z:\HPCHEM1\BNA_E\Data\BE103116\
 Data File : BE092488.D
 Acq On : 31 Oct 2016 20:59
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 01 11:17:38 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE102516.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 25 15:25:33 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	137	0.00
2	1,4-Dioxane	0.654	0.555	15.1	123	0.00
3	n-Nitrosodimethylamine	0.600	0.721	-20.2	172#	-0.04
4 S	2-Fluorophenol	0.758	1.261	-66.4#	250#	-0.01
5 S	Phenol-d6	1.060	1.873	-76.7#	259#	-0.02
6	bis(2-Chloroethyl)ether	1.251	1.447	-15.7	168#	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	141	0.00
8 S	Nitrobenzene-d5	0.281	0.344	-22.4	191#	-0.02
9	Naphthalene	1.081	1.061	1.9	138	0.00
10	Hexachlorobutadiene	0.166	0.169	-1.8	142	0.00
11	2-Methylnaphthalene	0.698	0.686	1.7	144	-0.02
12 I	Acenaphthene-d10	1.000	1.000	0.0	130	0.00
13 S	2,4,6-Tribromophenol	0.098	0.156	-59.2#	223#	-0.01
14 S	2-Fluorobiphenyl	1.218	1.292	-6.1	141	-0.01
15	Acenaphthylene	7.149	7.374	-3.1	139	0.00
16	Acenaphthene	1.141	1.152	-1.0	132	0.00
17	Fluorene	1.366	1.417	-3.7	140	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	117	0.00
19	Hexachlorobenzene	0.224	0.238	-6.2	130	0.00
20	Pentachlorophenol	0.039	0.107	-174.4#	382#	-0.02
21	Phenanthrene	1.292	1.277	1.2	127	0.00
22	Anthracene	1.163	1.311	-12.7	150#	-0.02
23	Fluoranthene	5.122	6.125	-19.6	149	0.00
24 I	Chrysene-d12	1.000	1.000	0.0	122	0.00
25	Pyrene	4.358	4.948	-13.5	149	-0.02
26 S	Terphenyl-d14	0.619	0.703	-13.6	150	-0.02
27	Benzo(a)anthracene	4.626	5.447	-17.7	152#	0.00
28	Chrysene	4.862	4.114	15.4	107	0.00
29	Bis(2-ethylhexyl)phthalate	0.385	0.713	-85.2#	298#	0.00
30	Indeno(1,2,3-cd)pyrene	4.286	4.970	-16.0	153#	0.00
31 I	Perylene-d12	1.000	1.000	0.0	126	0.01
32	Benzo(b)fluoranthene	1.245	1.234	0.9	144	0.00
33	Benzo(k)fluoranthene	1.234	1.362	-10.4	134	0.00
34 C	Benzo(a)pyrene	1.127	1.239	-9.9	157#	0.00
35	Dibenzo(a,h)anthracene	1.053	1.177	-11.8	155#	0.00
36	Benzo(g,h,i)perylene	4.540	4.863	-7.1	142	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0