

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE103116\
 Data File : BE092488.D
 Acq On : 31 Oct 2016 20:59
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4EC

Manual Integrations
 APPROVED

umangi
 11/1/2016 4:15:42 PM

Quant Time: Nov 01 11:17:38 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE102516.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 25 15:25:33 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.17	152	13931	5.00	ng	0.00
7) Naphthalene-d8	10.95	136	69050	5.00	ng	0.00
12) Acenaphthene-d10	14.82	164	44604	5.00	ng	0.00
18) Phenanthrene-d10	17.58	188	79331	5.00	ng	0.00
24) Chrysene-d12	21.75	240	100991	5.00	ng	0.00
31) Perylene-d12	24.12	264	86544	5.00	ng	0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.68	112	1405	0.62	ng	-0.01
5) Phenol-d6	7.34	99	2087	0.63	ng	-0.02
8) Nitrobenzene-d5	9.34	82	1903	0.55	ng	-0.02
13) 2,4,6-Tribromophenol	16.31	330	556	0.58	ng	-0.01
14) 2-Fluorobiphenyl	13.43	172	4612	0.42	ng	-0.01
26) Terphenyl-d14	20.17	244	5682	0.45	ng	-0.02
Target Compounds						Qvalue
2) 1,4-Dioxane	3.41	88	618	0.32	ng	98
3) n-Nitrosodimethylamine	3.77	42	803	0.53	ng	95
6) bis(2-Chloroethyl)ether	7.59	93	1613	0.50	ng	# 28
9) Naphthalene	10.99	128	5862	0.39	ng	97
10) Hexachlorobutadiene	11.28	225	935	0.41	ng	99
11) 2-Methylnaphthalene	12.62	142	3789	0.39	ng	94
15) Acenaphthylene	14.53	152	26314	0.41	ng	100
16) Acenaphthene	14.88	154	4112	0.40	ng	99
17) Fluorene	15.87	166	5057	0.42	ng	98
19) Hexachlorobenzene	16.89	284	1509	0.43	ng	# 65
20) Pentachlorophenol	17.23	266	676	0.93	ng	# 86
21) Phenanthrene	17.60	178	8107	0.40	ng	# 95
22) Anthracene	17.70	178	8321	0.45	ng	98
23) Fluoranthene	19.61	202	38871	0.48	ng	93
25) Pyrene	19.97	202	39977	0.45	ng	96
27) Benzo(a)anthracene	21.72	228	44005m	0.47	ng	
28) Chrysene	21.77	228	33235m	0.34	ng	
29) Bis(2-ethylhexyl)phthalate	21.63	149	5761	0.72	ng	# 74
30) Indeno(1,2,3-cd)pyrene	26.62	276	40156	0.46	ng	97
32) Benzo(b)fluoranthene	23.38	252	8545	0.40	ng	# 89
33) Benzo(k)fluoranthene	23.43	252	9427	0.44	ng	# 82
34) Benzo(a)pyrene	24.01	252	8576	0.44	ng	# 89
35) Dibenzo(a,h)anthracene	26.65	278	8150	0.45	ng	# 92
36) Benzo(g,h,i)perylene	27.40	276	33667	0.43	ng	# 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA_E\DATA\BE103116\
Data File : BE092488.D
Acq On : 31 Oct 2016 20:59
Operator : UM/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC

Manual Integrations
APPROVED

umangi
11/1/2016 4:15:42 PM

Quant Time: Nov 01 11:17:38 2016
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE102516.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 25 15:25:33 2016
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 8.17 min Scan# 576

Delta R.T. -0.00 min

Lab File: BE092488.D

Acq: 31 Oct 2016 20:59

Instrument :

BNA_E





ClientSampleId :

SSTDCCC0.4EC

Tgt Ion	Ratio	Lower	Upper
152	100		
150	153.7	106.0	159.0
115	56.0	31.2	46.8

Manual Integrations
APPROVED

umangi
11/1/2016 4:15:42 PM



#2

1,4-Dioxane

Concen: 0.32 ng




RT: 3.41 min Scan# 71


Delta R.T. -0.00 min

Lab File: BE092488.D

Acq: 31 Oct 2016 20:59

Tgt Ion	Ratio	Lower	Upper
88	100		
58	79.4	63.6	95.4
43	38.3	28.0	42.0





#3





n-Nitrosodimethylamine
Concen: 0.53 ng
RT: 3.77 min Scan# 102
Delta R.T. -0.04 min
Lab File: BE092488.D
Acq: 31 Oct 2016 20:59

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion: 42 Resp: 803
Ion Ratio Lower Upper
42 100
74 112.3 86.0 129.0
44 7.2 5.1 7.7

Manual Integrations
APPROVED




umangi
11/1/2016 4:15:42 PM




#4

2-Fluorophenol
Concen: 0.62 ng
RT: 5.68 min Scan# 364
Delta R.T. -0.01 min
Lab File: BE092488.D
Acq: 31 Oct 2016 20:59

Tgt Ion: 112 Resp: 1405
Ion Ratio Lower Upper
112 100
64 68.1 53.5 80.3
63 36.1 27.9 41.9





#5

Phenol-d6

Concen: 0.63 ng

RT: 7.34 min Scan# 540

Delta R.T. -0.02 min

Lab File: BE092488.D

Acq: 31 Oct 2016 20:59

Instrument :

BNA_E





ClientSampleId :

SSTDCCC0.4EC

Tgt Ion:	99	Resp:	2087
Ion	Ratio	Lower	Upper
99	100		
42	23.6	16.0	24.0
71	37.1	30.6	45.8

Manual Integrations
APPROVED

umangi
11/1/2016 4:15:42 PM



#6

bis(2-Chloroethyl)ether

Concen: 0.50 ng




RT: 7.59 min Scan# 551


Delta R.T. -0.00 min

Lab File: BE092488.D

Acq: 31 Oct 2016 20:59

Tgt Ion:	93	Resp:	1613
Ion	Ratio	Lower	Upper
93	100		
63	0.0	71.6	107.4#
95	30.6	24.0	36.0









#7
Naphthalene-d8
Concen: 5.00 ng
RT: 10.95 min Scan# 708
Delta R.T. -0.00 min
Lab File: BE092488.D
Acq: 31 Oct 2016 20:59

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion	Ratio	Lower	Upper
136	100		
137	11.6	8.2	12.4
54	5.7	9.6	14.4#
68	4.6	6.7	10.1#




Manual Integrations
APPROVED


umangi
11/1/2016 4:15:42 PM



#8
Nitrobenzene-d5
Concen: 0.55 ng
RT: 9.34 min Scan# 627
Delta R.T. -0.02 min
Lab File: BE092488.D
Acq: 31 Oct 2016 20:59

Tgt Ion	Ratio	Lower	Upper
82	100		
128	43.2	39.0	58.4
54	52.0	44.8	67.2





#9

Naphthalene

Concen: 0.39 ng

RT: 10.99 min Scan# 710

Delta R.T. -0.00 min

Lab File: BE092488.D

Acq: 31 Oct 2016 20:59

Instrument :

BNA_E





ClientSampleId :

SSTDCCC0.4EC

Tgt Ion:	128	Resp:	5862
Ion	Ratio	Lower	Upper
128	100		
129	12.0	11.2	16.8
127	15.0	11.6	17.4

Manual Integrations
APPROVED

umangi
11/1/2016 4:15:42 PM



#10

Hexachlorobutadiene

Concen: 0.41 ng




RT: 11.28 min Scan# 725


Delta R.T. -0.00 min

Lab File: BE092488.D

Acq: 31 Oct 2016 20:59

Tgt Ion:	225	Resp:	935
Ion	Ratio	Lower	Upper
225	100		
223	62.8	50.6	76.0
227	63.2	51.4	77.0





#11

2-Methylnaphthalene

Concen: 0.39 ng

RT: 12.62 min Scan# 814

Delta R.T. -0.02 min

Lab File: BE092488.D

Acq: 31 Oct 2016 20:59

Instrument :

BNA_E





ClientSampleId :

SSTDCCC0.4EC

Tgt Ion:	142	Resp:	3789
Ion Ratio	Lower	Upper	
142	100		
141	87.5	73.7	110.5
115	37.0	34.0	51.0

Manual Integrations
APPROVED

umangi
11/1/2016 4:15:42 PM



#12

Acenaphthene-d10

Concen: 5.00 ng




RT: 14.82 min Scan# 980


Delta R.T. -0.00 min

Lab File: BE092488.D

Acq: 31 Oct 2016 20:59

Tgt Ion:	164	Resp:	44604
Ion Ratio	Lower	Upper	
164	100		
162	104.4	71.4	107.0
160	45.2	27.4	41.2#









#13
 2,4,6-Tribromophenol
 Concen: 0.58 ng
 RT: 16.31 min Scan# 1088
 Delta R.T. -0.01 min
 Lab File: BE092488.D
 Acq: 31 Oct 2016 20:59

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4EC

Tgt Ion: 330 Resp: 556
 Ion Ratio Lower Upper
 330 100
 332 102.2 75.4 113.0
 141 44.1 31.0 46.6




Manual Integrations
 APPROVED


umangi
 11/1/2016 4:15:42 PM



#14
 2-Fluorobiphenyl
 Concen: 0.42 ng
 RT: 13.43 min Scan# 884
 Delta R.T. -0.01 min
 Lab File: BE092488.D
 Acq: 31 Oct 2016 20:59

Tgt Ion: 172 Resp: 4612
 Ion Ratio Lower Upper
 172 100
 171 35.6 30.1 45.1
 170 24.5 21.8 32.6





#15

Acenaphthylene

Concen: 0.41 ng

RT: 14.53 min Scan# 963

Delta R.T. -0.00 min

Lab File: BE092488.D

Acq: 31 Oct 2016 20:59

Instrument :

BNA_E





ClientSampleId :

SSTDCCC0.4EC

Tgt Ion:	152	Resp:	26314
Ion Ratio	Lower	Upper	
152	100		
151	5.0	3.9	5.9
153	13.1	10.4	15.6

Manual Integrations
APPROVED

umangi
11/1/2016 4:15:42 PM



#16

Acenaphthene

Concen: 0.40 ng




RT: 14.88 min Scan# 984


Delta R.T. -0.00 min

Lab File: BE092488.D

Acq: 31 Oct 2016 20:59

Tgt Ion:	154	Resp:	4112
Ion Ratio	Lower	Upper	
154	100		
153	435.5	350.8	526.2
152	214.9	171.1	256.7









#17
Fluorene
 Concen: 0.42 ng
 RT: 15.87 min Scan# 1050
 Delta R.T. -0.00 min
 Lab File: BE092488.D
 Acq: 31 Oct 2016 20:59

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4EC

Tgt Ion	Ratio	Lower	Upper
166	100		
165	99.5	78.2	117.4
167	13.4	11.0	16.4




Manual Integrations
 APPROVED


umangi
 11/1/2016 4:15:42 PM



#18
Phenanthrene-d10
 Concen: 5.00 ng
 RT: 17.58 min Scan# 1154
 Delta R.T. 0.00 min
 Lab File: BE092488.D
 Acq: 31 Oct 2016 20:59

Tgt Ion	Ratio	Lower	Upper
188	100		
94	2.9	3.2	4.8#
80	2.4	2.6	3.8#









#19
Hexachlorobenzene
Concen: 0.43 ng
RT: 16.89 min Scan# 1124
Delta R.T. 0.00 min
Lab File: BE092488.D
Acq: 31 Oct 2016 20:59

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion	Ratio	Lower	Upper
284	100		
142	0.0	29.1	43.7#
249	30.2	27.9	41.9




Manual Integrations
APPROVED


umangi
11/1/2016 4:15:42 PM



#20
Pentachlorophenol
Concen: 0.93 ng
RT: 17.23 min Scan# 1139
Delta R.T. -0.02 min
Lab File: BE092488.D
Acq: 31 Oct 2016 20:59

Tgt Ion	Ratio	Lower	Upper
266	100		
268	60.6	62.5	93.7#
264	77.8	57.2	85.8









#21
Phenanthrene
Concen: 0.40 ng
RT: 17.60 min Scan# 1155
Delta R.T. 0.00 min
Lab File: BE092488.D
Acq: 31 Oct 2016 20:59

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion	Ratio	Lower	Upper
178	100		
176	19.9	15.8	23.8
179	15.8	16.0	24.0




Manual Integrations
APPROVED


umangi
11/1/2016 4:15:42 PM



#22
Anthracene
Concen: 0.45 ng
RT: 17.70 min Scan# 1159
Delta R.T. -0.02 min
Lab File: BE092488.D
Acq: 31 Oct 2016 20:59

Tgt Ion	Ratio	Lower	Upper
178	100		
176	19.6	15.0	22.4
179	14.2	12.2	18.2









#23
Fluoranthene
 Concen: 0.48 ng
 RT: 19.61 min Scan# 1266
 Delta R.T. 0.00 min
 Lab File: BE092488.D
 Acq: 31 Oct 2016 20:59

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4EC

Tgt Ion	Ratio	Resp Lower	Resp Upper
202	100		
101	2.8	2.2	3.4
203	20.4	13.7	20.5




Manual Integrations
 APPROVED


umangi
 11/1/2016 4:15:42 PM



#24
Chrysene-d12
 Concen: 5.00 ng
 RT: 21.75 min Scan# 1375
 Delta R.T. 0.00 min
 Lab File: BE092488.D
 Acq: 31 Oct 2016 20:59

Tgt Ion	Ratio	Resp Lower	Resp Upper
240	100		
120	2.7	2.6	4.0
236	27.8	24.6	37.0









#25
 Pyrene
 Concen: 0.45 ng
 RT: 19.97 min Scan# 1287
 Delta R.T. -0.02 min
 Lab File: BE092488.D
 Acq: 31 Oct 2016 20:59

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4EC

Tgt Ion	Ratio	Lower	Upper
202	100		
200	5.4	4.2	6.4
203	19.7	13.9	20.9




Manual Integrations
 APPROVED


umangi
 11/1/2016 4:15:42 PM



#26
 Terphenyl-d14
 Concen: 0.45 ng
 RT: 20.17 min Scan# 1299
 Delta R.T. -0.02 min
 Lab File: BE092488.D
 Acq: 31 Oct 2016 20:59

Tgt Ion	Ratio	Lower	Upper
244	100		
212	11.4	6.7	10.1#
122	18.4	3.6	5.4#









#27
Benzo(a)anthracene
Concen: 0.47 ng m
RT: 21.72 min Scan# 1374
Delta R.T. 0.00 min
Lab File: BE092488.D
Acq: 31 Oct 2016 20:59

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion	Ratio	Lower	Upper
228	100		
226	26.8	23.6	35.4
229	19.2	13.3	19.9




Manual Integrations
APPROVED


umangi
11/1/2016 4:15:42 PM



#28
Chrysene
Concen: 0.34 ng m
RT: 21.77 min Scan# 1376
Delta R.T. 0.00 min
Lab File: BE092488.D
Acq: 31 Oct 2016 20:59

Tgt Ion	Ratio	Lower	Upper
228	100		
226	46.7	36.3	54.5
229	17.0	10.6	16.0









#29
 Bis(2-ethylhexyl)phthalate
 Concen: 0.72 ng
 RT: 21.63 min Scan# 1370
 Delta R.T. 0.00 min
 Lab File: BE092488.D
 Acq: 31 Oct 2016 20:59

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4EC

Tgt Ion	Ratio	Lower	Upper
149	100		
167	23.6	16.0	24.0
279	0.0	16.0	24.0




Manual Integrations
 APPROVED


umangi
 11/1/2016 4:15:42 PM



#30
 Indeno(1,2,3-cd)pyrene
 Concen: 0.46 ng
 RT: 26.62 min Scan# 1717
 Delta R.T. -0.00 min
 Lab File: BE092488.D
 Acq: 31 Oct 2016 20:59

Tgt Ion	Ratio	Lower	Upper
276	100		
138	27.9	20.3	30.5
277	24.5	19.7	29.5









#31
 Perylene-d12
 Concen: 5.00 ng
 RT: 24.12 min Scan# 1547
 Delta R.T. 0.01 min
 Lab File: BE092488.D
 Acq: 31 Oct 2016 20:59

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4EC

Tgt Ion	Ratio	Lower	Upper
264	100		
260	24.8	20.1	30.1
265	22.4	17.9	26.9




Manual Integrations
 APPROVED


umangi
 11/1/2016 4:15:42 PM



#32
 Benzo(b)fluoranthene
 Concen: 0.40 ng
 RT: 23.38 min Scan# 1483
 Delta R.T. -0.00 min
 Lab File: BE092488.D
 Acq: 31 Oct 2016 20:59

Tgt Ion	Ratio	Lower	Upper
252	100		
253	26.5	18.7	28.1
125	20.4	10.5	15.7









#33
Benzo(k)fluoranthene
Concen: 0.44 ng
RT: 23.43 min Scan# 1487
Delta R.T. -0.00 min
Lab File: BE092488.D
Acq: 31 Oct 2016 20:59

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion	Ratio	Lower	Upper
252	100		
253	26.7	20.3	30.5
125	31.2	9.6	14.4#




Manual Integrations
APPROVED


umangi
11/1/2016 4:15:42 PM



#34
Benzo(a)pyrene
Concen: 0.44 ng
RT: 24.01 min Scan# 1537
Delta R.T. -0.00 min
Lab File: BE092488.D
Acq: 31 Oct 2016 20:59

Tgt Ion	Ratio	Lower	Upper
252	100		
253	27.0	19.0	28.6
125	22.9	11.8	17.8#









#35
Dibenzo(a,h)anthracene
Concen: 0.45 ng
RT: 26.65 min Scan# 1719
Delta R.T. -0.00 min
Lab File: BE092488.D
Acq: 31 Oct 2016 20:59

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion	Ratio	Lower	Upper
278	100		
139	23.0	13.8	20.8#
279	29.3	21.4	32.2

Manual Integrations
APPROVED

umangi
11/1/2016 4:15:42 PM



#36
Benzo(g,h,i)perylene
Concen: 0.43 ng
RT: 27.40 min Scan# 1763
Delta R.T. -0.00 min
Lab File: BE092488.D
Acq: 31 Oct 2016 20:59

Tgt Ion	Ratio	Lower	Upper
276	100		
277	28.2	19.8	29.8
138	30.7	19.8	29.6#

