




Data Path : Z:\HPCHEM1\BNA E\Data\BE103117\
Data File : BE094671.D
Acq On : 31 Oct 2017 23:01
Operator : SJ/JU
Sample : I5844-08
Misc :
ALS Vial : 19 Sample Multiplier: 1

Quant Time: Nov 01 02:59:10 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE102517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 01 02:55:07 2017
Response via : Initial Calibration



TIC: BE094671.D

(10) Phenanthrene-d10 (I)

17.257min (-0.000) 0.40ng/ul m

response 9413

Ion	Exp%	Act%
188.00	100	100
94.00	11.00	6.72#
80.00	11.80	5.83#
0.00	0.00	0.00