




Data Path : Z:\HPCHEM1\BNA E\Data\BE103117\
Data File : Be094656.d
Acq On : 31 Oct 2017 13:46
Operator : SJ/JU
Sample : PB103511BL
Misc :
ALS Vial : 5 Sample Multiplier: 1

Quant Time: Nov 01 01:56:59 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE102517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 01 01:43:26 2017
Response via : Initial Calibration



TIC: Be094656.d

(20) Perylene-d12 (I)

23.960min (-0.008) 0.40ng/ul m

response 12350

Ion	Exp%	Act%
264.00	100	100
260.00	30.10	24.12
265.00	103.50	27.53#
0.00	0.00	0.00