

Data Path : Z:\HPCHEM1\BNA E\DATA\BE110117\
 Data File : BE094687.D
 Acq On : 1 Nov 2017 12:39
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 01 14:09:30 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE102317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 23 22:40:47 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	107	0.00
2	1,4-Dioxane	0.677	0.654	3.4	90	-0.01
3	n-Nitrosodimethylamine	0.648	0.481	25.8#	79	0.01
4 S	2-Fluorophenol	1.370	1.110	19.0	87	0.04
5 S	Phenol-d6	2.069	1.949	5.8	104	0.04
6	bis(2-Chloroethyl)ether	1.600	1.385	13.4	94	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	113	0.00
8 S	Nitrobenzene-d5	0.319	0.270	15.4	99	0.02
9	Naphthalene	4.527	4.001	11.6	97	0.02
10	Hexachlorobutadiene	0.164	0.160	2.4	108	0.00
11 SURR	2-Methylnaphthalene-d10	0.642	0.552	14.0	95	0.00
12	2-Methylnaphthalene	0.768	0.652	15.1	95	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	101	0.01
14 S	2,4,6-Tribromophenol	0.110	0.092	16.4	88	0.03
15 S	2-Fluorobiphenyl	1.410	1.309	7.2	93	0.01
16	Acenaphthylene	2.112	1.829	13.4	88	0.00
17	Acenaphthene	1.356	1.215	10.4	89	0.00
18	Fluorene	1.738	1.495	14.0	87	0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	95	0.03
20	4-Bromophenyl-phenylether	0.218	0.176	19.3	79	0.03
21	Hexachlorobenzene	0.191	0.278	-45.5#	121	0.00
22	Pentachlorophenol	0.024	0.026	-8.3	102	0.06
23	Phenanthrene	1.141	1.437	-25.9#	119	0.00
24	Anthracene	1.176	1.115	5.2	89	0.00
25 SURR	Fluoranthene-d10	4.676	5.179	-10.8	92	0.00
26	Fluoranthene	1.420	1.541	-8.5	90	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	100	0.00
28	Pvrene	1.399	1.258	10.1	91	0.00
29 S	Terphenyl-d14	0.754	0.666	11.7	90	0.00
30	Benzo(a)anthracene	1.302	1.173	9.9	90	0.01
31	Chrysene	1.322	1.208	8.6	94	0.01
32	Bis(2-ethylhexyl)phthalate	3.193	2.808	12.1	96	-0.01
33	Indeno(1,2,3-cd)pyrene	1.222	0.909	25.6#	75	0.05
34 I	Perylene-d12	1.000	1.000	0.0	97	0.02
35	Benzo(b)fluoranthene	1.294	1.123	13.2	87	0.01
36	Benzo(k)fluoranthene	1.379	1.266	8.2	90	0.02
37 C	Benzo(a)pyrene	1.255	1.168	6.9	91	0.02
38	Dibenzo(a,h)anthracene	1.127	0.908	19.4	80	0.03
39	Benzo(a,h,i)perylene	1.055	0.748	29.1#	70	0.06

Data Path : Z:\HPCHEM1\BNA E\DATA\BE110117\
Data File : BE094687.D
Acq On : 1 Nov 2017 12:39
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Nov 01 14:09:30 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE102317.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 23 22:40:47 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			