

Data Path : Z:\HPCHEM1\BNA_E\Data\BE110217\
 Data File : BE094699.D
 Acq On : 2 Nov 2017 17:27
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 02 18:21:11 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE102317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 23 22:40:47 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	115	0.00
2	1,4-Dioxane	0.677	0.649	4.1	95	-0.01
3	n-Nitrosodimethylamine	0.648	0.473	27.0#	84	0.01
4 S	2-Fluorophenol	1.370	1.018	25.7#	85	0.05
5 S	Phenol-d6	2.069	1.856	10.3	106	0.04
6	bis(2-Chloroethyl)ether	1.600	1.410	11.9	102	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	121	0.00
8 S	Nitrobenzene-d5	0.319	0.269	15.7	106	0.03
9	Naphthalene	4.527	3.981	12.1	103	0.02
10	Hexachlorobutadiene	0.164	0.156	4.9	114	0.00
11 SURR	2-Methylnaphthalene-d10	0.642	0.555	13.6	102	0.00
12	2-Methylnaphthalene	0.768	0.659	14.2	103	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	108	0.01
14 S	2,4,6-Tribromophenol	0.110	0.072	34.5#	74	0.03
15 S	2-Fluorobiphenyl	1.410	1.284	8.9	97	0.01
16	Acenaphthylene	2.112	1.815	14.1	94	0.00
17	Acenaphthene	1.356	1.212	10.6	96	0.00
18	Fluorene	1.738	1.484	14.6	93	0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	102	0.03
20	4-Bromophenyl-phenylether	0.218	0.179	17.9	86	0.03
21	Hexachlorobenzene	0.191	0.273	-42.9#	127	0.00
22	Pentachlorophenol	0.024	0.020	16.7	83	0.07
23	Phenanthrene	1.141	1.317	-15.4	118	0.00
24	Anthracene	1.176	1.161	1.3	99	0.00
25 SURR	Fluoranthene-d10	4.676	5.033	-7.6	96	0.00
26	Fluoranthene	1.420	1.506	-6.1	95	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	105	0.00
28	Pyrene	1.399	1.245	11.0	95	0.00
29 S	Terphenyl-d14	0.754	0.658	12.7	93	0.00
30	Benzo(a)anthracene	1.302	1.139	12.5	92	0.01
31	Chrysene	1.322	1.207	8.7	99	0.01
32	Bis(2-ethylhexyl)phthalate	3.193	2.644	17.2	95	-0.01
33	Indeno(1,2,3-cd)pyrene	1.222	0.827	32.3#	72	0.05
34 I	Perylene-d12	1.000	1.000	0.0	99	0.01
35	Benzo(b)fluoranthene	1.294	1.162	10.2	91	0.00
36	Benzo(k)fluoranthene	1.379	1.373	0.4	99	0.01
37 C	Benzo(a)pyrene	1.255	1.170	6.8	93	0.02
38	Dibenzo(a,h)anthracene	1.127	0.860	23.7	77	0.03
39	Benzo(g,h,i)perylene	1.055	0.701	33.6#	67	0.06

Data Path : Z:\HPCHEM1\BNA_E\Data\BE110217\
Data File : BE094699.D
Acq On : 2 Nov 2017 17:27
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Nov 02 18:21:11 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE102317.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 23 22:40:47 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			