

Data Path : Z:\HPCHEM1\BNA E\Data\Be110515\
 Data File : BE091062.D
 Acq On : 5 Nov 2015 9:49
 Operator : UM/NP
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.485

Quant Time: Nov 06 05:39:11 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM02.2-EPA-SIM-BE102715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 28 03:57:22 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	113	0.00
2 SURR	1,4-Dioxane-d8	0.288	0.250	13.2	92	-0.03
3	1,4-Dioxane	0.325	0.286	12.0	97	-0.03
4 I	Naphthalene-d8	1.000	1.000	0.0	133	0.00
5	Naphthalene	0.986	0.923	6.4	121	0.00
6 SURR	2-Methylnaphthalene-d10	0.546	0.540	1.1	131	-0.01
7	2-Methylnaphthalene	0.643	0.631	1.9	129	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	141	0.00
9	Acenaphthylene	1.998	1.917	4.1	131	0.00
10 C	Acenaphthene	1.368	1.284	6.1	129	0.00
11	Fluorene	1.757	1.533	12.7	114	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	107	0.00
13	Pentachlorophenol	0.038	0.039	-2.6	137	0.00
14	Phenanthrene	4.619	4.398	4.8	101	0.00
15	Anthracene	4.200	4.296	-2.3	113	0.01
16 SURR	Fluoranthene-d10	1.127	1.000	11.3	101	0.00
17 C	Fluoranthene	5.406	4.759	12.0	98	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	101	0.00
19	Pyrene	4.544	4.589	-1.0	101	0.00
20	Benzo(a)anthracene	1.027	1.041	-1.4	105	0.00
21	Chrysene	1.183	1.100	7.0	93	0.00
22 I	Perylene-d12	1.000	1.000	0.0	113	0.00
23	Benzo(b)fluoranthene	1.249	1.150	7.9	102	0.00
24	Benzo(k)fluoranthene	1.281	1.099	14.2	93	0.00
25 C	Benzo(a)pyrene	1.150	1.108	3.7	107	0.00
26	Indeno(1,2,3-cd)pyrene	4.847	4.393	9.4	100	0.00
27	Dibenzo(a,h)anthracene	1.155	1.084	6.1	106	0.00
28	Benzo(q,h,i)perylene	5.083	4.575	10.0	101	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0