

Data Path : Z:\HPCHEM1\BNA E\DATA\BE110515\
 Data File : BE091075.D
 Acq On : 5 Nov 2015 20:16
 Operator : UM/NP
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.486

Quant Time: Nov 06 05:45:22 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM02.2-EPA-SIM-BE102715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 06 05:41:57 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	74	0.00
2 SURR	1,4-Dioxane-d8	0.288	0.245	14.9	59	-0.03
3	1,4-Dioxane	0.325	0.270	16.9	60	-0.03
4 I	Naphthalene-d8	1.000	1.000	0.0	87	0.00
5	Naphthalene	0.986	0.940	4.7	80	0.00
6 SURR	2-Methylnaphthalene-d10	0.546	0.550	-0.7	87	-0.01
7	2-Methylnaphthalene	0.643	0.651	-1.2	87	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	94	0.00
9	Acenaphthylene	1.998	1.936	3.1	89	0.00
10 C	Acenaphthene	1.368	1.306	4.5	88	0.00
11	Fluorene	1.757	1.522	13.4	76	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	74	0.00
13	Pentachlorophenol	0.038	0.040	-5.3	98	0.00
14	Phenanthrene	4.619	4.357	5.7	69	0.00
15	Anthracene	4.200	4.154	1.1	75	0.00
16 SURR	Fluoranthene-d10	1.127	1.024	9.1	72	0.00
17 C	Fluoranthene	5.406	4.881	9.7	70	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	74	0.00
19	Pyrene	4.544	4.559	-0.3	73	0.00
20	Benzo(a)anthracene	1.027	1.070	-4.2	79	0.00
21	Chrysene	1.183	1.091	7.8	67	0.00
22 I	Perylene-d12	1.000	1.000	0.0	85	0.00
23	Benzo(b)fluoranthene	1.249	1.095	12.3	73	0.00
24	Benzo(k)fluoranthene	1.281	1.172	8.5	74	0.00
25 C	Benzo(a)pyrene	1.150	1.088	5.4	79	0.00
26	Indeno(1,2,3-cd)pyrene	4.847	4.422	8.8	75	0.00
27	Dibenzo(a,h)anthracene	1.155	1.099	4.8	80	0.00
28	Benzo(q,h,i)perylene	5.083	4.634	8.8	77	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0