

Data Path : Z:\HPCHEM1\BNA E\DATA\BE110515\
 Data File : BE091078.D
 Acq On : 5 Nov 2015 22:45
 Operator : UM/NP
 Sample : SSTDC00.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.487

Quant Time: Nov 06 06:43:24 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM02.2-EPA-SIM-BE102715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 06 06:42:14 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)

1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	75	-0.01
2 SURR	1,4-Dioxane-d8	0.288	0.241	16.3	59	-0.03
3	1,4-Dioxane	0.325	0.271	16.6	61	-0.03
4 I	Naphthalene-d8	1.000	1.000	0.0	87	0.00
5	Naphthalene	0.986	0.954	3.2	82	0.00
6 SURR	2-Methylnaphthalene-d10	0.546	0.557	-2.0	89	-0.01
7	2-Methylnaphthalene	0.643	0.651	-1.2	87	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	93	0.00
9	Acenaphthylene	1.998	1.944	2.7	88	0.00
10 C	Acenaphthene	1.368	1.318	3.7	88	0.00
11	Fluorene	1.757	1.551	11.7	77	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	74	0.00
13	Pentachlorophenol	0.038	0.040	-5.3	97	0.00
14	Phenanthrene	4.619	4.360	5.6	69	0.00
15	Anthracene	4.200	4.239	-0.9	77	0.00
16 SURR	Fluoranthene-d10	1.127	1.038	7.9	72	0.00
17 C	Fluoranthene	5.406	4.927	8.9	70	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	74	0.00
19	Pyrene	4.544	4.485	1.3	72	0.00
20	Benzo(a)anthracene	1.027	1.082	-5.4	80	0.00
21	Chrysene	1.183	1.086	8.2	67	0.00
22 I	Perylene-d12	1.000	1.000	0.0	85	0.00
23	Benzo(b)fluoranthene	1.249	1.179	5.6	79	0.00
24	Benzo(k)fluoranthene	1.281	1.104	13.8	70	0.00
25 C	Benzo(a)pyrene	1.150	1.111	3.4	81	0.00
26	Indeno(1,2,3-cd)pyrene	4.847	5.403	-11.5	92	0.00
27	Dibenzo(a,h)anthracene	1.155	1.108	4.1	81	0.00
28	Benzo(q,h,i)perylene	5.083	4.593	9.6	76	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0