

Data Path : Z:\HPCHEM1\BNA_E\Data\BE110615\
 Data File : BE091080.D
 Acq On : 6 Nov 2015 8:09
 Operator : UM/NP
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.488

Quant Time: Nov 09 13:08:12 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM02.2-EPA-SIM-BE102715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 09 05:49:45 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	77	0.00
2 SURR	1,4-Dioxane-d8	0.288	0.248	13.9	62	0.00
3	1,4-Dioxane	0.325	0.284	12.6	66	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	92	0.00
5	Naphthalene	0.986	0.939	4.8	85	0.00
6 SURR	2-Methylnaphthalene-d10	0.546	0.546	0.0	91	0.00
7	2-Methylnaphthalene	0.643	0.640	0.5	90	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	97	0.00
9	Acenaphthylene	1.998	1.932	3.3	91	0.00
10 C	Acenaphthene	1.368	1.291	5.6	89	0.00
11	Fluorene	1.757	1.553	11.6	79	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	78	0.00
13	Pentachlorophenol	0.038	0.035	7.9	89	0.00
14	Phenanthrene	4.619	4.332	6.2	72	0.00
15	Anthracene	4.200	4.170	0.7	80	0.00
16 SURR	Fluoranthene-d10	1.127	1.008	10.6	74	0.00
17 C	Fluoranthene	5.406	4.897	9.4	73	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	76	0.00
19	Pyrene	4.544	4.478	1.5	74	0.00
20	Benzo(a)anthracene	1.027	1.044	-1.7	80	0.00
21	Chrysene	1.183	1.109	6.3	71	0.00
22 I	Perylene-d12	1.000	1.000	0.0	87	0.00
23	Benzo(b)fluoranthene	1.249	1.167	6.6	80	0.00
24	Benzo(k)fluoranthene	1.281	1.105	13.7	72	0.00
25 C	Benzo(a)pyrene	1.150	1.071	6.9	80	0.00
26	Indeno(1,2,3-cd)pyrene	4.847	4.445	8.3	77	0.00
27	Dibenzo(a,h)anthracene	1.155	1.097	5.0	82	0.00
28	Benzo(g,h,i)perylene	5.083	4.560	10.3	77	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0