

Data Path : Z:\HPCHEM1\BNA_E\Data\BE110615\
 Data File : BE091094.D
 Acq On : 6 Nov 2015 17:12
 Operator : UM/NP
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.489

Quant Time: Nov 09 14:01:43 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM02.2-EPA-SIM-BE102715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 09 05:49:45 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	73	0.00
2 SURR	1,4-Dioxane-d8	0.288	0.255	11.5	61	0.00
3	1,4-Dioxane	0.325	0.293	9.8	64	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	88	0.00
5	Naphthalene	0.986	0.943	4.4	81	0.00
6 SURR	2-Methylnaphthalene-d10	0.546	0.538	1.5	86	0.00
7	2-Methylnaphthalene	0.643	0.622	3.3	84	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	91	0.00
9	Acenaphthylene	1.998	1.957	2.1	86	0.00
10 C	Acenaphthene	1.368	1.289	5.8	84	0.00
11	Fluorene	1.757	1.549	11.8	75	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	74	0.00
13	Pentachlorophenol	0.038	0.030	21.1#	73	0.00
14	Phenanthrene	4.619	4.272	7.5	68	0.00
15	Anthracene	4.200	4.015	4.4	73	0.00
16 SURR	Fluoranthene-d10	1.127	0.968	14.1	68	0.00
17 C	Fluoranthene	5.406	4.730	12.5	68	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	70	0.00
19	Pyrene	4.544	4.458	1.9	68	0.00
20	Benzo(a)anthracene	1.027	1.018	0.9	71	0.00
21	Chrysene	1.183	1.130	4.5	66	0.00
22 I	Perylene-d12	1.000	1.000	0.0	80	0.00
23	Benzo(b)fluoranthene	1.249	1.152	7.8	72	0.00
24	Benzo(k)fluoranthene	1.281	1.113	13.1	66	0.00
25 C	Benzo(a)pyrene	1.150	1.073	6.7	73	0.00
26	Indeno(1,2,3-cd)pyrene	4.847	4.311	11.1	69	0.00
27	Dibenzo(a,h)anthracene	1.155	1.085	6.1	74	0.00
28	Benzo(g,h,i)perylene	5.083	4.603	9.4	71	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0